



FG-84096 FastGene® Mag Virus Kit (1-96)

FG-845K FastGene® Mag Virus Kit (5000)

Table of Contents

1.	General Information	3
1.1	Intended Use	3
1.2	Kit specifications	3
1.3	Basic principle.....	3
2.	Materials	4
2.1	Kit Contents	4
2.2	Reagents, consumables and equipment to be supplied by the user	5
3.	Kit usage	6
3.1	Storage Conditions	6
3.2	Preparation of reagents.....	6
3.3	Safety instructions.....	6
3.4	Considerations.....	7
3.5	Magnetic Separation systems	7
3.6	Shaker settings	7
3.7	Product use limitations	8
4.	Protocols	9
4.1	Sample materials and pre-treatment procedures.....	9
4.2	Manual processing of 200 µL liquid or pre-treated samples.....	10
4.3	Protocol for the KingFisher Flex™ Magnetic Particle Processor.....	11
4.3.1	KingFisher BindIt software protocol	11
4.3.2	Preparation of processing plates	12
4.3.3	Detailed instructions	13
5.	Troubleshooting	14
6.	Ordering Information	15
7.	Contact Information	15

1. General Information

1.1 *Intended Use*

FastGene® Mag Virus Kit is for Research Use Only (RUO). The kit is suited for qualified personnel only.

The kit is intended for manual and automated isolation of nucleic acids (DNA and RNA) from a wide range of samples. Processing time for the preparation of 96 samples is about 40 minutes. The kit requires no phenol/chloroform extraction or alcohol precipitation and eliminates the need for repeated centrifugation, vacuum filtration or column separation. It allows safe handling of potentially infectious samples and is designed to avoid sample-to-sample cross-contaminations. The obtained nucleic acids can be used directly as template for downstream applications such as PCR, qPCR, qRT-PCR or any kind of enzymatic reaction.

FastGene® Mag Virus Kit is suitable for use with blood samples, liquid samples (e.g. plasma, serum, urine, swab washes), tissue samples, feces. For details on the individual procedures for sample pre-treatment see below.

FastGene® Mag Virus Kit magnetic beads are optimized for use in isolating total nucleic acids. The beads are easy to handle and are supplied in an optimized storage buffer for increased suspension time.

1.2 *Kit specifications*

The kit provides reagents for extraction of total nucleic acids from 200 µL liquid sample or 200 µL homogenized tissue samples, cells or suspended feces. Total nucleic acids are finally eluted in a volume of 100 µL Elution Buffer.






The obtained nucleic acids should be used for qPCR, qRT-PCR immediately after extraction. Storage at <-20°C is recommended for later analysis.

1.3 *Basic principle*

Samples are lysed under denaturing conditions by adding Lysis Buffer MVL and Proteinase K. During the incubation the released nucleic acids can bind to the FastGene® Mag Beads magnetic beads. After magnetic separation and discard of the supernatant, the beads are washed three times using alcoholic buffers (Wash Buffers MVWI and MVWII) to remove contaminants and salts. A drying step makes sure all traces of ethanol from the final wash steps are removed. Finally, purified nucleic acids are eluted with low-salt Elution Buffer and can directly be used for downstream applications.

2. Materials

2.1 Kit Contents

	96 preps FG-84096	5000 preps FG-845K
Lysis Buffer MVL 	20 mL	1 L
Binding Buffer MVB 	40 mL	2 x 1 L
Wash Buffer MVWI 	2 x 80 mL	8 x 1 L
Wash Buffer MVWII 	80 mL	4 x 1 L
Elution Buffer MVE 	20 mL	1 L
Proteinase K	20 mg (for 1 mL working solution)	5 x 200 g (for 50 mL working solution)
Poly-A-RNA	0.3 mg	15 mg
Poly-A-RNA Buffer	0.5 mL	20 mL
FastGene® Mag Beads	2 mL	100 mL

2.2 Reagents, consumables and equipment to be supplied by the user

- molecular biology grade (nuclease free) water to reconstitute Proteinase K

Consumables/equipment:

Protocol	Manual use	Automated use
Sample containers	1.5 or 2 mL microtubes	Recommended: Riplate®SW 96, PP, 2 mL, (Ritter, 43001-0020) Nunc™ 96-well Polypropylene DeepWell™ Storage Plate 2.0 mL, (Thermo Scientific, Art.No. 278752)
Magnetic separation	MM-Separator for 1.5 or 2 mL tubes	MM-Separator for 96 DeepWell Plates
Final container	1.5 or 2 mL microtubes	96-well microplate
Mixing	Tube Vortexer	Microplate shaker (min. 1000 RPM)

3. Kit usage

3.1 Storage Conditions

All kit components including **Proteinase K** (lyophilized), **Poly-A-RNA** (lyophilized) and **FastGene® Mag Beads** can be stored at room temperature. Store ready solutions of Proteinase K and Poly-A-RNA in suitable aliquots at -20°C. When stored under the conditions mentioned, the kit is stable as indicated by the expiry date on the label.

3.2 Preparation of reagents

- Reconstitute Proteinase K:
FG-84096 (96 preps), add 1 mL of **molecular biology grade water** to the vial of **Proteinase K** and vortex to dissolve. Store solution of Proteinase K in aliquots at -20°C.

FG-845K (5000 preps), add 10 mL of **molecular biology grade water** to each vial of **Proteinase K** and vortex to dissolve. Store solution of Proteinase K in aliquots at -20°C.
- Reconstitute Poly-A-RNA:
Add 120 µL (or 6 mL for FG-845K) of **Poly-A-RNA Buffer** to each vial of **Poly-A-RNA** and vortex to dissolve. Store solutions of Poly-A-RNA in aliquots at -20°C.
- If there is any precipitate present in the buffers, warm the buffer to 25-37°C to dissolve the precipitate before use.
- Immediately before use, resuspend the FastGene® Mag Beads by vortexing for 20 seconds.
- Samples should be thoroughly mixed before use.

3.3 Safety instructions

Take appropriate safety measures, such as wearing a suitable lab coat, disposable gloves, and protective goggles. Follow local legal requirements for working with biological materials.

More information is found in the safety data sheets (SDS), available at www.nippongenetics.eu.

Infectious potential of liquid waste left over after using the FastGene® Mag Virus Kit was not tested. Even though contamination of waste with residual infectious material is unlikely, it cannot be excluded completely. Therefore, liquid waste should be handled as being potentially infectious, and discarded according to local safety regulations.

3.4 Considerations

1. To avoid cross-contamination and DNA degradation, change pipette tips after each use and use nuclease-free filter-tips.
2. Avoid leaving bottles open to prevent contamination or evaporation of the kit reagents.
3. Do not combine components of different kits unless the lot numbers are identical.
4. Process only as many samples in parallel as the magnetic separator allows.
5. The elution can be done in smaller volumes of Elution Buffer MVE. Although this may result in higher nucleic acids concentrations, overall yield may be lower. The yield may also be increased by prolonging the incubation time, and with pre-heated Elution Buffer (56°C).
6. The Elution Buffer does not contain EDTA.
7. Avoid samples containing coagulates or precipitates, as this may result in poor results or quality. Centrifuge samples before use.
8. The kit is compatible with whole blood treated with EDTA and citrate. Heparin is co-isolated and may interfere with subsequent PCR analyses.
9. It may occur that a small amount of beads is accidentally transferred with the final DNA sample, but most likely this will not inhibit subsequent applications. However, if desired another separation step can be performed to remove the beads.

3.5 Magnetic Separation systems

FastGene® Mag Virus Kit has been designed for use on the MM-Separator for 1.5 and 2 mL tubes or MM-Separator for 96 DeepWell plates.

For use with other magnetic separators, please contact the technical support info@nippongenetics.de.

FastGene® Mag Virus Kit is compatible with KingFisher™ Flex Magnetic Particle Processor by Thermo Scientific™. Protocols and consumables are available on request.

3.6 Shaker settings

The speed settings for the microplate shaker described in the protocols that follow were defined with a specific instrument and microplate. When first using a plate shaker for

incubation steps, the speed settings have to be set carefully for each specific plate to prevent cross contamination and spillage. Setting the speed of the shaker can be done by loading a microplate with a volume of dyed water equal to the working volume during each step, and step-wise increasing the shaker speed until droplets are observed on the surface of the plate. Set the shaker speed lower again.

3.7 Product use limitations

FastGene® Mag Virus Kit is intended for **Research Use Only**. Do not use for other purposes than intended. The kit components can be used only once.

No guarantee is offered when using sample material other than specified. It is recommended to check the suitability of the purified nucleic acids for each selected qPCR / qRT-PCR assay.

The end-user has to validate the performance of the kit for any particular use, since the performance characteristics of the kits have not been validated for any specific application. FastGene® kits may be used in clinical diagnostic laboratory systems after the laboratory has validated the complete diagnostic system as required by CLIA' 88 regulations in the U.S. or equivalents in other countries.

The product is intended for use by trained personnel. The isolated nucleic acids can be used in most genomic applications, such as PCR, qPCR.

Diagnostic results generated using the sample preparation procedure should only be interpreted with regard to other clinical or laboratory findings. Adequate controls should be used in each set of isolations, especially when used for diagnostic purposes.

4. Protocols

4.1 Sample materials and pre-treatment procedures

Preparation of the Lysis Master Mix

For each sample mix 200 μ L Lysis Buffer MVL with 1 μ L reconstituted Poly-A-RNA and 10 μ L of reconstituted Proteinase K solution (20 mg/mL). Prepare an excess of the Lysis Master Mix to compensate for pipetting inaccuracy especially when using multichannel pipettes etc. Use the Lysis Master Mix immediately after preparation.

Recommendations for sample pre-treatment

Sample material	Pre-treatment procedure	Further processing
Liquid samples (e.g. plasma, serum, blood, swab wash solutions)	n/a	Use up to 200 μ L of the liquid sample and add 211 μ L Lysis Master Mix
Dry swabs without transport media	Add an appropriate volume of molecular biology grade water to dried swab (swab should be completely submerged) and shake/vortex vigorously for 15 minutes. Take 200 μ L for further processing	Use 200 μ L of the solution and add 211 μ L Lysis Master Mix
Feces / stool	Add 1.5 mL molecular biology grade water to a pea-size amount of feces. Mix well by vortexing. Spin down at low g-forces to remove remaining particulate sample residuals. Use 200 μ L of the suspension for further processing.	Use up to 200 μ L of the centrifuged suspension and add 211 μ L Lysis Master Mix
Tissue samples	Mechanically homogenize <30 mg of tissue sample in 500 μ L molecular biology grade water using suitable devices (NG010_Tissue Grinder Mixy Professional). Spin down for 1 min at 8,000 x g to remove debris. Use 200 μ L of the suspension for further processing.	Use up to 200 μ L of the centrifuged suspension and add 211 μ L Lysis Master Mix

For samples not mentioned in the table above please contact Nippon Genetics Europe GmbH for support protocols.

4.2 Manual processing of 200 μ L liquid or pre-treated samples

Before starting:

- Check if Lysis Master Mix was prepared according to section 4.1.
- Pre-treat of samples (if required) according to section 4.1
- Vortex magnetic beads thoroughly into a homogeneous suspension.

This protocol is intended for manual use of the kit. It can also be used as a guideline to set up an automated procedure on liquid handling instruments. For this suitable 96-well plates and accessories can be used. Make sure that the liquid handling devices is equipped with the required devices (shaker, incubator, magnetic separator etc.).

1. Transfer **200 μ L sample** into a microtube. If the volume is lower than 200 μ L, bring the volume up to 200 μ L with 1 x PBS buffer or molecular biology grade water.
2. Add **211 μ L Lysis Master Mix**. Mix immediately and incubate on a shaker for **10 min with shaking at 1000 RPM**. Spin down briefly to collect any sample from the microtube lid.
3. Add **20 μ L FastGene[®] Mag Beads** and **400 μ L Binding Buffer MVB** ●. Mix and incubate on a shaker for **5 min with shaking at 1000 RPM**. Spin down briefly to collect any sample from the microtube lid.
4. Place the samples on the magnetic separator and wait at least 1 minute to collect the beads. Remove supernatants without disturbing the attracted magnetic bead pellet.
5. Remove the sample plate from the magnetic separator and add **800 μ L Wash Buffer MVWI** ● to the tubes. Resuspend the beads by incubation of the samples on a shaker for **1 min at 1000 RPM** (alternatively pipette the wash buffer up and down until the beads are resuspended completely). Place the samples in a magnetic separator and wait at least 1 minute to collect the beads. Remove the supernatants without disturbing the attracted bead pellet.
6. Repeat step 5 one more time with **800 μ L Wash Buffer MVWI** ● and one more time with **800 μ L Wash Buffer MVWII** ●.
7. Dry the beads on air for **10 min** to evaporate the ethanol completely. If necessary, briefly spin down and remove any buffer residues before drying the attracted magnetic beads.

8. Remove the samples from the magnetic separator and add **100 µL Elution Buffer MVE ●**. Resuspend the attracted magnetic beads by repeated pipetting up and down. **Incubate** on a shaker for **10 min at 1000 RPM**.
9. Place the tubes in a magnetic separator and wait at least 1 minute to collect the beads. Transfer the eluates to new tubes. The purified nucleic acids in the eluate are now ready to use.
 - *If the transferred eluates appear turbid, briefly centrifuge the samples and carefully transfer the eluates.*
 - *If the transferred eluates contain magnetic particles, place the tubes on the magnetic separator again, separate for 1 minute and transfer the eluates.*
 - *The nucleic acids can be eluted with a lower volume of Elution Buffer (depending on the expected yield). The minimum volume for elution is 30 µL and this can reduce the yield. If a large amount of nucleic acids is expected, the volume of Elution Buffer can be increased.*

4.3 Protocol for the KingFisher Flex™ Magnetic Particle Processor

4.3.1 KingFisher BindIt software protocol

Please contact Nippon Genetics Europe GmbH for the most recent BindIt software method files. We provide the corresponding files for direct upload on the KingFisher™ magnetic particle processors. A PDF description of the method file is included. Refer to the BindIt software manual regarding the upload procedure of the supplied software files to the instrument.

4.3.2 Preparation of processing plates

Initial plate filling for instrument set-up:





Plate	Type*)	Reagent	Volume
Sample Plate	2 mL Deepwell Plate with square wells for KingFisher™	(pre-treated) sample adjusted to 200 µL	200 µL
		Lysis Master Mix	211 µL
		FastGene® Mag Beads	20 µL
		Binding Buffer MVB ●	400 µL
Wash Buffer I 1st	2 mL Deepwell Plate with square wells for KingFisher™	Wash Buffer MVWI ●	800 µL
Wash Buffer I 2nd	2 mL Deepwell Plate with square wells for KingFisher™	Wash Buffer MVWI ●	800 µL
Wash Buffer II	2 mL Deepwell Plate with square wells for KingFisher™	Wash Buffer MVWII ●	800 µL
Elution Buffer	200 µL square-well Elution Plate for KingFisher™	Elution Buffer MVE ●	100 µL
Tip plate	2 mL Deepwell Plate with square wells for KingFisher™	Empty, for loading Tip-Comb only	N/A

To be added after the initial lysis step on the KingFisher™ Flex magnetic particle processor

*) We strongly recommend using only the plates which are intended to use on the KingFisher™ magnetic particle processor. Using unsuitable plates may result in extraction failure or instrument damage.

4.3.3 Detailed instructions

Follow exactly the instructions as given below. **Do not change the order of reagent addition for the Sample Plate.** Label all plates thoroughly and unambiguously to avoid any misloading during the instrument loading procedure.

1. Prepare samples according to section 4.1. Add **200 µL of the sample** to each well of the Sample Plate. If the volume is lower than 200 µL, bring the volume up to 200 µL with 1 x PBS buffer or molecular biology grade water.
2. Add **211 µL of Lysis Master Mix** to each well of the Sample Plate. Mix by pipetting up and down until a homogeneous mixture is visible (typically 3 – 4 times). Incubate on a suitable plate shaker for 10 min with shaking at 1000 rpm. Continue with preparing the remaining plates before continuing with the sample plate.
3. Prepare two plates for the 1st and 2nd wash steps with **Wash Buffer MVWI**. Add **800 µL Wash Buffer MVWI**  to each well of the corresponding deep-well plates.
4. Prepare one plates for the 3rd wash step with **Wash Buffer MVWII**. Add **800 µL Wash Buffer MVWII**  to each well of the corresponding deep-well plate.
5. Prepare one plate for **Elution Buffer**. Add **100 µL Elution Buffer MVE**  to each well of the corresponding plate.
6. Add **20 µL of FastGene® Mag Beads** to each well of the Sample Plate.
7. Add **400 µL of Binding Buffer MVB**  to each well of the Sample Plate.
8. Switch on the KingFisher™ Flex magnetic particle processor and select the “**FastGene® Mag Virus**” protocol from the user defined protocols
9. Start the protocol.
10. Load the plates to the instrument, following the instructions on the instrument display. Order of plates start with the tip plate and ends with the sample plate. The purification process starts immediately after loading the sample plate to the instrument.

Make sure that all plates are inserted in the same orientation (especially when using partially filled plates). Place the A1 well of each plate to the A1 mark on the instruments turntable.

11. At the end of the method remove all plates from the instrument. Follow the information on the instrument display.

5. Troubleshooting

Problem	Possible causes	Comments and suggestions
Low nucleic acid yield	Insufficient sample lysis	- Optimize sample pre-treatment, make sure that Lysis Master Mix includes Proteinase K
	Inefficient binding to the magnetic particles	- Use correct amount of all reagents - Increase mixing steps incubation time for binding step - Mix sample during lysis / binding incubation
	Insufficient washing procedure	- Make sure that beads are completely resuspended in the washing steps.
	Incomplete elution	- Drying of Wash Buffer MVWII may have been incomplete - Completely resuspend the beads in the elution step.
Problems in downstream applications / contamination in DNA sample	Ethanol in the eluted DNA	- Increase the drying time to 15 minutes
	Salt in the eluate	- Make sure that all supernatants are properly removed. - Avoid carry-over of Lysis Master Mix, Binding Buffer MVB or Wash Buffers to the eluate.
	Magnetic beads remaining in the eluate	- Place the tubes with eluates in the magnetic separator again, and transfer the supernatant to a new container. High amounts of co-purified genomic DNA (e.g. for cell and tissue samples) may cause high viscosity of the eluate and force incomplete bead separation in the elution step. Use higher volume of elution buffer and/or reduce sample input.

6. Ordering Information

Cat. No.	Product	Content
FG-84096	FastGene® Mag Virus Kit	96 preparations
FG-845K	FastGene® Mag Virus Kit	5000 preparations
FG-82050	FastGene® RNA Viral Kit	50 preparations
FG-82300	FastGene® RNA Viral Kit	300 preparations
FG-SSMAG2	FastGene® Magnastand 0.2	
FG-SSMAG1.5	FastGene® Magnastand 1.5	
FG-SSMAG96	FastGene® Magnastand 96	
NG010	Tissue Grinder Mixy Professional	1
NG006	Pestles for Tissue Grinder Mixy Professional	100

7. Contact Information

EUROPE

Nippon Genetics Europe GmbH
 Mariaweilerstraße 28-30
 D-52349 Dueren
 Germany
 Phone: +49 (0) 2421 55496-0
 Email: info@nippongenetics.eu
info@nippongenetics.de

For more detailed product information, contact details, questions, or trouble shooting please visit our English website www.nippongenetics.eu or our German website www.nippongenetics.eu/de.

JAPAN

Nippon Genetics Co., Ltd
 Koraku Mori Bldg., 18F
 1-4-14 Koraku, Bunkyo-ku, Tokyo
 112-0004 Japan
 Phone: +81 (3) 3813 0961
 Email: info@genetics-n.co.jp

For more detailed product information, contact details, questions, or trouble shooting please visit our Japanese website www.n-genetics.co.

FastGene® is a registered trademark of Nippon Genetics Europe GmbH.