

## Technical File

1	Classification of Directive 93/42/EEC	Product group	Dermatological Instrument
		Short name of product	Biopsy punches
		Product name	Biopsy punch
		Sterilization method	Gamma sterilization
		Classification	Class II a (Rule 6, Subclause 1, No Indent)
2	Declaration of conformity	See attached Declaration of conformity "Biopsy Punches" issued on 2016-08-31	
3	General description of the product	<p>A) Stainless steel tube blade with polypropylene plastic handle. Each in blister pack and 20 pcs. in a dispenser box, 50 boxes per export carton.</p> <p>B) Stainless steel tube blade with polypropylene plastic handle with plunger. Each in blister pack and 20 pcs. in a dispenser box, 50 boxes per export carton.</p> <p>Dimension of knife : –See attached Declaration of conformity "Biopsy Punches" issued on 2016-08-31</p>	
4	Essential requirements checklist	Recorded in the design history file. MDD Annex I – KCFI02A-F-7, Rev.13, control number 25-02, Rev.7 issued on 2016-08-31	
5	Applied standards	See attached Declaration of conformity "Biopsy Punches" issued on 2016-08-31	
6	Intended use	Used for cutting, dissecting and/or filing of tissue mostly skin of human's body This product is single use.	
7	Function	This product has a handle, and a blade. This product is Used for cutting, dissecting and/or filing of tissue mostly skin of human's body.	
8	Accessories and detachable parts	N/A	
9	Material	Standards and components of materials each parts - Manufacturing control standard of the biopsy punch [KH-M-BP140]	
10	Mechanical drawings	<p>Blades: SM-TB002-K-02 issued on 2016-04-22</p> <p>Handle: SM-TH013-K-01 issued on 2015-06-25</p> <p>SM-TH001-K-01 issued on 2001-10-10</p> <p>951018-01 issued on 1995-10-18</p> <p>SM-TH002-K-01 issued on 2001-10-10</p> <p>SM-TH004-K-13 issued on 2015-06-24</p> <p>Plunger: SM-TP005-K-08 issued on 2006-04-27</p> <p>SM-TP004-K-05 issued on 2006-04-27</p> <p>Spring: SM-TP002-K-02 issued on 2005-02-21</p> <p>Products: SM-TQ001-K-01 issued on 2003-03-17</p> <p>SM-TQ002-K-02 issued on 2015-06-25</p>	
11	Label and instructions for use	Examination of label [M15602+QQ2T], [M16202+QQ2T] Biopsy punch primary package 1406B, 1602B, 1602C Biopsy punch secondary package 1406B, 1602B	
12	Packaging	<p>Seal peel test [M15331=KQ2T]</p> <p>Sterility test [M13Z26=PQ2T]</p> <p>Drop test [M13Y11=FQ2T]</p> <p>Microbial barrier test [M15827=KQ4T]</p> <p>Transportation test [M15609=PF2T]</p>	

13	Manufacturing process	Process management regulations [KQFI05A] Processing conditions table [NO.80C]、[NO.80D] Manufacturing flow chart - Manufacturing control standard of the biopsy punch [KH-M-BP-010]
14	Inspection and quality assurance techniques	Manufacturing control standard of the biopsy punch [KH-M-BP]
15	S h e l f l i f e	5 years after sterilization
16	Sterilization validation	Procedure for gamma sterilization validation of ISO 11137-2 Method Vdmax <sup>25</sup> [KQKH65A] Report for gamma sterilization validation of biopsy punches [20150804] Report for gamma sterilization validation of biopsy punche with plunger [20160823]
17	E O G residuals	N/A
18	Mechanical tests	Resistance test [M15825=KQ3T]、[M16822=KQ2A] Sharpness test [M16224=KQ3T] Visual and functional test [M16224=FQ2T]、[M16822=FQ2A]
19	Risk evaluation	Summary of Risk evaluation for Biopsy punch [KQFI11A-F4 No.05 issued on 2016-08-31]
20	Clinical evaluation	Summary of Clinical evaluation for Biopsy punch [KQKH23A-F2 No.05 issued on 2016-08-31]
21	Biocompatibility	Summary of Biocompatibility for Biopsy punch [KQKH64A-F2 No.05 issued on 2016-08-31]
22	U s a b i l i t y	Summary of Usability engineering for Biopsy punch [KQKH61A-F4 No.05 issued on 2016-08-31]

Approval	Create
	

Revision History – 1 / 1

Revision No.	Prepared on	Matter for Revision	Reason for Revision	Create by
1	2015/09/03	N/A	Create New	Shinoda
2	2016/03/04	Section : 2,3,4,5,11,12,16,18,19,22	Revision of sections	Shinoda
3	2016/08/31	Section : 2,3,4,5,10,16,18,19,20,21,22	Revision of sections	Shinoda