

Standard Analysis Certificate

PRODUCT :CONDUCTIVITY STANDARD 5 μ S/cm RS
CODE :575000
METHOD :10447

TEST	U.M.	SPECIFICATION
Description	-	Clear colourless liquid
Identification	-	Positive
Conductivity at 25°C	μ S/cm	4.95 \div 5.05

Traceable to Standard NIST N° 999a
Confidence interval: \pm 0.02 μ S/cm

DOMINIQUE DUTSCHER SAS

Date

:11/01/2020

QUALITY CONTROL RESPONSIBLE
B. COULANGE (VDR)