

# Standard Analysis Certificate

**PRODUCT** :CONDUCTIVITY STANDARD 20 µS/cm RS  
**CODE** :575020  
**METHOD** :10449

TEST	U.M.	SPECIFICATION
Description	-	Clear colourless liquid
Identification	-	Positive
Conductivity at 25°C	µS/cm	19.80 ÷ 20.20

Traceable to Standard NIST N° 999a  
 Confidence interval: ± 0.1 µS/cm

DOMINIQUE DUTSCHER SAS

Date :11/01/2020

**QUALITY CONTROL RESPONSIBLE**  
**B. COULANGE (VDR)**