

Standard Analysis Certificate

PRODUCT :CONDUCTIVITY STANDARD 100 µS/cm RS
CODE :575050
METHOD :10452

TEST	U.M.	SPECIFICATION
Description	-	Clear colourless liquid
Identification	-	Positive
Conductivity at 25°C	µS/cm	99.0 ÷ 101.0

Traceable to Standard NIST N° 999a
Confidence interval: ± 0.1 µS/cm

DOMINIQUE DUTSCHER SAS

Date

:11/01/2020

QUALITY CONTROL RESPONSIBLE
B. COULANGE (VDR)