

Standard Analysis Certificate

PRODUCT :CONDUCTIVITY STANDARD 147 µS/cm RS
CODE :575060
METHOD :10453

TEST	U.M.	SPECIFICATION
Description	-	Clear colourless liquid
Identification	-	Positive
Conductivity at 25°C	µS/cm	145.5 ÷ 148.5

Traceable to Standard NIST N° 999a
 Confidence interval: ± 0.1 µS/cm

DOMINIQUE DUTSCHER SAS

Date :11/01/2020

QUALITY CONTROL RESPONSIBLE
B. COULANGE (VDR)