

Standard Analysis Certificate

PRODUCT :STANDARD PER CONDUTTIMETRIA 1000 μ S/cm RS
CODE :575090
METHOD :10456

TEST	U.M.	SPECIFICATION
Description	-	Clear colourless liquid
Identification	-	Positive
Conductivity at 25°C	μ S/cm	990.0 \div 1010.0

Traceable to Standard NIST N° 999a
 Confidence interval: $\pm 1 \mu$ S/cm

DOMINIQUE DUTSCHER SAS

Date :11/01/2020

QUALITY CONTROL RESPONSIBLE
B. COULANGE (VDR)