

# Standard Analysis Certificate

**PRODUCT** :CONDUCTIVITY STANDARD 1413  $\mu\text{S}/\text{cmRS}$   
**CODE** :575100  
**METHOD** :10457

TEST	U.M.	SPECIFICATION
Description	-	Clear colourless liquid
Identification	-	Positive
Conductivity at 25°C	$\mu\text{S}/\text{cm}$	1399 $\div$ 1427

Traceable to Standard NIST N° 999a  
 Confidence interval:  $\pm 1 \mu\text{S}/\text{cm}$

DOMINIQUE DUTSCHER SAS

Date :11/01/2020

**QUALITY CONTROL RESPONSIBLE**  
**B. COULANGE (VDR)**