

# Standard Analysis Certificate

**PRODUCT** :CONDUCTIVITY STANDARD 5000  $\mu$ S/cm RS  
**CODE** :575110  
**METHOD** :10458

TEST	U.M.	SPECIFICATION
Description	-	Clear colourless liquid
Identification	-	Positive
Conductivity at 25°C	$\mu$ S/cm	4950 $\div$ 5050

Traceable to Standard NIST N° 999a  
 Confidence interval:  $\pm 20 \mu$ S/cm

DOMINIQUE DUTSCHER SAS

Date :11/01/2020

**QUALITY CONTROL RESPONSIBLE**  
**B. COULANGE (VDR)**