

Standard Analysis Certificate

PRODUCT :CONDUCTIVITY STANDARD 12880 μ S/cm RS
CODE :575130
METHOD :10460

TEST	U.M.	SPECIFICATION
Description	-	Clear colourless liquid
Identification	-	Positive
Conductivity at 25°C	μ S/cm	12751 \div 13009

Traceable to Standard NIST N° 999a
 Confidence interval: \pm 20 μ S/cm

DOMINIQUE DUTSCHER SAS

Date :11/01/2020

QUALITY CONTROL RESPONSIBLE
B. COULANGE (VDR)