

Declaration of Conformity



Name and address of the company

Bio Molecular Systems Pty Ltd
5/3 Northward St
Upper Coomera, QLD, 4209
Australia

We herewith declare that the product

Myra

Meets all applicable requirements of the following European Directives

Electromagnetic Compatibility Directive (EMC)	2014/30/EU
Restriction of Hazardous Substances Directive (RoHS)	2011/65/EU
Waste Electrical and Electronic Equipment Directive (WEEE)	2012/19/EU

and the relevant harmonised standards

EN 61326-1:2013 Electrical equipment for measurement, control and laboratory use- EMC requirements – Part 1 General requirements.

EN 61000-4-2:2009 Electromagnetic compatibility (EMC)- Part 4-2: Testing and measurement techniques- Electrostatic discharge immunity test.


EN 61000-4-3:2010 Electromagnetic compatibility (EMC)- Part 4-3: Testing and measurement techniques- Radiated, radio-frequency, electromagnetic field immunity test.

EN 61000-4-4:2010 Electromagnetic compatibility (EMC)-Part 4-4: Testing and measurement techniques- Electrical fast transient/burst immunity test.

EN 61000-4-5:2009 Electromagnetic compatibility (EMC)- Part 4-5: Testing and measurement techniques-Surge immunity test.

EN 61000-4-6:2009 Electromagnetic compatibility (EMC)- Part 4-6: Testing and measurement techniques- Immunity to conducted disturbances, induced by radio-frequency fields.

EN 61000-4-11:2004 Electromagnetic compatibility (EMC) Parts 4-11: Testing and measurement techniques – Voltage dips, short interruptions and voltage various immunity tests.


John Corbett

Director

22nd January 2019


bio molecular systems