CORNING

Certificate of compliance

CORNING GOSSELIN SAS
123 Route de Caestre
59190 - BORRE
FRANCE
www.corning.com/lifesciences
Refer to website for regional contact information

Date of certificate : 2026 - 05 - 26

Product description
Container 40mL PP Red Screw Cap, Sterile *100/Bag, 1000/Case*

Catalogue number
TP30C-002

Manufacturing date
2026/04/01

Lot number
972699

Quality Management System

Complies with the current version of the ISO9001 standard.

Sterility Assurance Level

Product has been sterilized and dosimetrically released per the requirements of ANSI/AAMI/ISO 11137, "Sterilization of health care products - Radiation". Products meet a minimum Sterility Assurance Level (SAL) of 10-3.

Expiration date
2029/04/01

Expiration statement

This product is designed to have an expiration date of 3 years from the date of manufacture regarding sterility.

Performance testing (AQL)

Inspection based on a statistically significant number of samples towards to internal or agreed specifications that determine Assurance Quality Level.

DOMINIQUE DUTSCHER SAS

This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release



Sandra Bianchi
Plant Quality Manager, Borre
BianchiS@corning.com