PDS No. 6279x5	PRODUCT DATA SHEET	Page 1 of 1
Revision 04	CELLview [™] , Cell Culture Dish, 35 x 10 mm, Glass Bottom, Advanced TC™, Sterile	6
	Greiner Item-No. 6279x5	greiner bio-one
Valid for Item-No.:	627965 (sterile) 627975 (sterile)	

1.	Description / Specification			
1.1	Description	Cell culture dish with glass bottom for microscopic applications, with vents,		
		sterile, Advanced TC™ surface		
		Glass bottom: maximal spectral transmission, no autofluorescence, no		
		depolarisation of light		
		627965: with one compartment,		
		627975: with four compartments		
1.2	Dimensions	See Customer Drawing		
		Glass Thickness: 175 μm (+/- 15 μm)		
1.3 Volume 627965: working volume: 5,0 ml		627965: working volume: 5,0 ml		
		627975: working volume: 0,1-0,5 ml / per compartment		
		627965: growth area: 8,7 cm ²		
		627975: growth area: 1,9 cm ² / per compartment		
1.4	Material / Resin	Dish: PS (Polystyrene), free of heavy metal		
		Bottom: borosilicate glass; hydrolytic class 1 (DIN-ISO 719)		
		Lid: PS (Polystyrene), free of heavy metal		
1.5	Colour	Dish: clear		
		Glass Bottom: high transparent		
		Lid: clear		
1.6	Sterilisation	SAL 10 ⁻³		
1.7	Quality Control	- Raw Material-Control: physical testing		
	-	- Product-Control: testing of attributive and variable characteristics in		
		accordance with the valid specification		
1.8	Other Information	- For single use only		
		- Manufactured according to ISO 8255-1:1986		

2.	Features	
2.1	Basic features	Free of detectable DNase/RNase, human DNA and pyrogens. Contents non-cytotoxic Adhesive: resistant to DMSO and 70% ethanol, compatible with cell culture media
2.2	Temperature range	+ 4℃ to +37℃
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	N/A
2.5	Chemical Resistance	See homepage: https://www.gbo.com/en_INT/know-how-services/download-center.html
2.6	Shelf life	2 years after month of production
2.7	Other Information	-

3.	Packaging	
3.1	Pieces / Bag	10
3.2	Pieces / Box	40
3.3	Lot-No.	E JJ MM XXX (manufacturing facility, year, month, consecutive SAP-No.)
3.4	Other Information	Certificate of Quality

4.	Other Information
4.1	Research use only. Not for diagnostics.

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this
Revision	Date	Date	Date	document or drawing is confidential and proprietory to Greiner Bio-One GmbH. This
03	2 March 2015	3 March 2015	3 March 2015	document may not be reproduced for any
Date	Name	Name	Name	reason without written permission from Greiner Bio-One GmbH. All rights of design, invention,
27.07.2011	S. Kaelberer	Dr. T.Schreiber	A. Schulz	and copyright are reserved.