PDS No. 6501xx	PRODUCT DATA SHEET				Page 1 of 1
Revision 08	96 Well Microplate, PS, U-Bottom				6
Revision 08	Greiner Item-No. 6501xx				greiner
Valid for Item-No.:	650101	<b>650160</b> (sterile)	<b>650161</b> (sterile)	<b>650162</b> (sterile	e) <b>650180</b> (sterile)
	<b>650185</b> (sterile)				

1.	Description / Specification	
1.1	Description	PS Plate, 96 well, clear, solid U-bottom, alphanumeric well coding
		650101: standard
		650160: physical surface treatment, sterile
		650161, -162: standard, sterile
		650180: physical surface treatment, with standard lid, sterile
		650185: hydrophobic surface (without treatment), with standard lid, sterile
1.2	Dimensions	See customer drawing
1.3	Volume per well	Total volume: 323 µl (mathematically calculated)
		Working volume: 40 - 280 μl
		650160, -180: growth area: 0,35 cm <sup>2</sup>
1.4	Material / Resin	PS (Polystyrene), free of heavy metal
1.5	Colour	Clear
1.6	Sterilization	650101: No
		650160, -161, -162, -180, -185: SAL 10 <sup>-3</sup>
1.7	Quality Control	- Raw Material-Control: physical testing
	·	- Product-Control: testing of attributive and variable characteristics in
		accordance with the valid specification
1.8	Intended Use	General laboratory product for the processing and storage of samples to be
		used by qualified personnel in a laboratory environment.
1.9	Other Information	For single use only

2.	Features			
2.1	Basic features	All Item-No.: Free of detectable DNase/RNase, human DNA and pyrogens. 650160, -180, -185: contents non-cytotoxic		
2.2	Temperature range	For application: -20°C to +60°C		
2.3	Autoclavability	No		
2.4	Centrifugation, max. RCF	1000 x g: swinging-bucket rotor		
2.5	Chemical Resistance	See homepage: https://www.gbo.com/en_INT/know-how-services/download-center.html		
2.6	Shelf life	650101: N/A 650161, -162, -185: 5 years 650160, -180: 4 years		
2.7	Other Information	-		

3.	Packaging	650101	650160	650161	650162	650180, -185
3.1	Pieces / Bag	10	1	2	25	1
3.2	Pieces / Box	100	75	100	375	60
3.3	Lot-No.	E YY MM XXX	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)			
3.4	Other Information	Certificate of C	Certificate of Quality to download			

4.	Other Information
	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this	
Revision	Date	Date	Date	document or drawing is confidential and proprietory to Greiner Bio-One GmbH. This	
07	8 April 2021	8 April 2021	8 April 2021	document may not be reproduced for any reason without written permission from Greiner	
Date 27.11.2014	Name S. Kaelberer	Name P. Wachter	Name A. Illig	Bio-One GmbH. All rights of design, invention and copyright are reserved.	