PDS No. 6531xx	PRODUCT DATA SHEET			Page 1 of 1	
Revision 03	60 Well Microtest Plate (Terasaki Plate), PS			6	
	Greiner Item-No. 6531xx				greiner bio-one
Valid for Item-No.:	653102	653180	653190		

1.	Description / Specification		
1.1	Description	PS Microtest Plate (Terasaki Plate), 60 Well, conical well profile, flat well bottom, alphanumeric well coding, with lid	
		653102: Microbatch under Oil Applications, standard 653180, -190: HLA / Microtest Plate, physical surface treatment	
1.2	Dimensions	See customer drawing	
1.3	Volume per well	10 μΙ	
1.4	Material / Resin	Plate and lid: PS (Polystyrene), free of heavy metal	
1.5	Colour	Plate and lid: clear	
1.6	Sterilization	No	
1.7	Quality Control	Raw Material-Control: physical and immunological testing Product-Control: testing of attributive and variable characteristics in accordance with the valid specification	
1.8	Other Information	For single use only	

2.	Features	
2.1	Basic features	-
2.2	Temperature range	-20°C to +60°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	N/A
2.5	Chemical Resistance	See homepage: https://www.gbo.com/en_INT/know-how-services/download-center.html
2.6	Shelf life	653102: n/a 653180, -190: 1 year after month of production
2.7	Other Information	-

3.	Packaging	653102	653180	653190	
3.1	Pieces / Bag	10 / bag	10 / bag	120 / folding carton	
3.2	Pieces / Box	580	270	480	
3.3	Lot-No.	E JJ MM XXX (manu	E JJ MM XXX (manufacturing facility, year, month, consecutive SAP-No.)		
3.4	Other Information	6 -7 /			

4.	Other Information

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this document or drawing is confidential and proprietory to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 02	Date 26 November 2014	Date 28 November 2014	Date 28 November 2014	
Date	Name	Name	Name	
18.12.2009	S. Kaelberer	Dr. U. Honisch	A. Schulz	