


PDS No. 6541xx	PRODUCT DATA SHEET			Page 1 of 1
Revision 03	72 Well Microtest Plate (Terasaki Plate), PS			
	Greiner Item-No. 6541xx			
Valid for Item-No.:	654102	654180		

1.	Description / Specification	
1.1	Description	PS Microtest Plate (Terasaki Plate), 72 Well, conical well profile, flat well bottom, alphanumeric well coding, with lid 654102: Microbatch under Oil Applications, standard 654180: HLA / Microtest Plate, physical surface treatment
1.2	Dimensions	See customer drawing
1.3	Volume per well	10 µl
1.4	Material / Resin	Plate and lid: PS (Polystyrene), free of heavy metal
1.5	Colour	Plate and lid: clear
1.6	Sterilization	No
1.7	Quality Control	- Raw Material-Control: physical and immunological testing - <u>Product-Control</u> : testing of attributive and variable characteristics in accordance with the valid specification
1.8	Other Information	For single use only

2.	Features	
2.1	Basic features	-
2.2	Temperature range	-20°C to +60°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	N/A
2.5	Chemical Resistance	See homepage: https://www.gbo.com/en_INT/know-how-services/download-center.html
2.6	Shelf life	654102: n/a 654180: 1 year after month of production
2.7	Other Information	-

3.	Packaging	
3.1	Pieces / Bag	10
3.2	Pieces / Box	270
3.3	Lot-No.	E JJ MM XXX (manufacturing facility, year, month, consecutive SAP-No.)
3.4	Other Information	-

4.	Other Information
	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 02	Date 26 November 2014	Date 28 November 2014	Date 28 November 2014	
Date 18.12.2009	Name S. Kaelberer	Name Dr. U. Honisch	Name A. Schulz	

DISCLAIMER: The description of a certain product can only be considered as a guidance, because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.