


PDS No. 655167 655169	PRODUCT DATA SHEET			Page 1 of 1
Revision 00	ThinCert® 96 Well, PS Receiver Plate, F-Bottom, with Lid, Sterile			
	Greiner Item-No. 65516x			
Valid for Item-No.:	655167	655169 (sterile)		

1.	Description / Specification	
1.1	Description	ThinCert® 96 Well, PS Receiver Plate, F-Bottom, high profile lid with condensation rings, sterile 655167: non-treated 655169: physical surface treatment
1.2	Dimensions	See Customer Drawing
1.3	Volume per well	Working volume: 120-300 µl
1.4	Material / Resin	Plate / lid: PS (Polystyrene), free of heavy metal
1.5	Colour	Plate / lid: clear
1.6	Sterilization	SAL 10 ⁻³
1.7	Quality Control	- <u>Raw Material-Control</u> : physical testing - <u>Product-Control</u> : testing of attributive and variable characteristics in accordance with the valid specification
1.8	Intended Use	655167: General laboratory product for the processing and storage of samples to be used by qualified personnel in a laboratory environment. 655169: General laboratory product for cell culture to be used by qualified personnel in a laboratory environment.
1.9	Other Information	For single use only

2.	Features	
2.1	Basic features	All Item-No.: free of detectable DNase/RNase, human DNA and pyrogens. Contents non-cytotoxic
2.2	Temperature range	For application: -20°C to +60°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	N/A
2.5	Chemical Resistance	See homepage: https://www.gbo.com/en_INT/know-how-services/download-center.html
2.6	Shelf life	655167: 5 years 655169: 4 years
2.7	Other Information	-

3.	Packaging	
3.1	Pieces / Bag	8
3.2	Pieces / Box	32
3.3	Lot-No.	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)
3.4	Other Information	Certificate of Quality to download

4.	Other Information
	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision -	Date 19 August 2022	Date 22 August 2022	Date 23 August 2022	
Date -	Name S. Kaelberer	Name P. Wachter	Name A. Illig	

DISCLAIMER: The description of a certain product can only be considered as a guidance because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.