PDS No. 6576xx	PRODUCT DATA SHEET				Page 1 of 1	
ThinCert™Cell Culture Inserts for				for		
Revision 04	Revision 04 6 Well Plates					
	Greiner Item-No. 6576xx				greiner bio-one	
Valid for Item-No.:	<b>657610</b> (sterile)	<b>657630</b> (sterile)	<b>657631</b> (sterile)	<b>657638</b> (sterile	) <b>657640</b> (sterile)	
	<b>657641</b> (sterile)	•		_		

1.	Description / Specification			
1.1	Description	ThinCert™ Cell Culture Inserts, hanging geometry, sealed capillary pore membrane, physical surface treatment, sterile		
		Pore size (µm)	Pore density (cm <sup>-2</sup> )	
		657610: 1.0	657610: 2 x 10 <sup>6</sup>	
		657630: 3,0	657630: 0,6 x 10 <sup>6</sup>	
		657631: 3,0	657631: 2 x 10 <sup>6</sup>	
		657638: 8,0	657638: 0,15 x 10 <sup>6</sup>	
		<i>657640</i> : 0,4	657640: 1 x 10 <sup>8</sup>	
		657641: 0,4	657641: 2 x 10 <sup>6</sup>	
1.2	Dimensions	See Customer Drawing		
		Culture surface: 452,4 mm <sup>2</sup>		
1.3	Volume	Working volume ThinCert™: 1,0 – 3,6 ml		
		Working volume well: 2,0 – 4,15 ml		
1.4	Material / Resin	ThinCert™ inserts: PS (Polystyrene), free of heavy metal		
		Capillary pore membrane: PET, free of heavy metal		
1.5	Colour	ThinCert™ inserts: clear		
		Capillary pore membrane: 657610, -630, -641: transparent		
	<u> </u>	657631, -638, -640: translucent		
1.6	Sterilisation	SAL 10 <sup>-3</sup>		
1.7	Quality Control	Product-Control: testing of attributive and variable characteristics in		
		accordance with the valid specification		
1.8	Other Information	USP class VI certified		

2.	Features	
2.1	Basic features	Free of detectable DNase/RNase, human DNA and pyrogens.
		Contents non-cytotoxic
2.2	Temperature range	-20℃ to +60℃
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	N/A
2.5	Chemical Resistance	See homepage:
		https://www.gbo.com/en_INT/know-how-services/download-center.html
2.6	Shelf life	5 years after month of production
2.7	Other Information	-

3.	Packaging	ThinCert™ inserts	Multiwell plates
3.1	Pieces (single packed; paper/plastic)	1	1
3.2	Pieces / Box	24	4
3.3	Lot-No.	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)	
3.4	Other Information	Certificate of Quality	

4.	Other Information
4.1	Research use only. Not for diagnostics.

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this
Revision 03	Date 20 April 2015	Date 21 April 2015	Date 21 April 2015	document or drawing is confidential and proprietory to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Date 02.12.2014	Name S. Kaelberer	Name Dr. S. Mühlfriedel	Name A. Schulz	