PDS No. 6732xx	PRODUCT DATA SHEET			Page 1 of 1	
Povision 05	Sapphire PCR 8-Tube Strips, 0.2 ml, without Cap			6	
Revision 05	Greiner Item-No. 6732xx				greiner bio-one
Valid for Item-No.:	673210	673271	673273	673274	673275
	673276	673277			

1.	Description / Specification			
1.1	Description	Sappphire PCR 8-tube strips, 0.2 ml, without cap		
		- Comprehensive thermal cycler compatibility		
		- Suitable for 8-cap strips		
1.2	Dimensions	See Customer Drawing		
1.3	Volume	<u>Tube</u> : 0.2 ml		
1.4	Material / Resin	PP (Polypropylene), free of heavy metal		
1.5	Colour	673210: clear		
		673271: assorted (red, blue, green, yellow, violet)		
		673273: red		
		673274: blue		
		673275: green		
		673276: yellow		
		673277: violet		
1.6	Sterilization	No		
1.7	Quality Control	ity Control Product-Control: testing of attributive and variable characteristics in		
	-	accordance with the valid specification		
1.8	Other Information	For single use only		

2.	Features		
2.1	Basic features	Free of detectable DNase/RNase, human DNA and pyrogens	
2.2	Temperature range	-20°C to +105°C	
2.3	Autoclavability	No	
2.4	Centrifugation, max. RCF	N/A	
2.5	Chemical Resistance	See homepage: https://www.gbo.com/en_INT/know-how-services/download-center.html	
2.6	Shelf life	5 years after month of production	
2.7	Other Information	-	

3.	Packaging	
3.1	Pieces / Box	125
3.2	Pieces / Case	1.250
3.3	Lot-No.	5-digit 12345 (digit 1 & 2 identify the year and 3 & 4 identify the week)
3.4	Other Information	Packaging labelled with product symbols for easier identification and warehousing

4.	Other Information
	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this
Revision	Date	Date	Date	document or drawing is confidential and proprietory to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
04	13 April 2016	30 June 2016	1 July 2016	
Date	Name	Name	Name	
22.04.2015	S. Kaelberer	Dr. R. Heller	Dr. A. Ganser	