


PDS No. 675001/ 675061	<b>PRODUCT DATA SHEET</b>			Page 1 of 1
Revision 03	96 Well ELISA Microplate, Solid Bottom, Half Area			 <b>greiner bio-one</b>
	Item-No. 675001 / 675061			
Valid for Item-No.:	<b>675001</b>	<b>675061</b>		

1.	Description / Specification	
1.1	Description	ELISA Microplate, 96 well, half area well profile, solid bottom, alphanumeric well coding 675001: Microlon <sup>®</sup> 200, medium binding 675061: Microlon <sup>®</sup> 600, high binding
1.2	Dimensions	See customer drawing
1.3	Volume	Total volume: 199 µl (mathematically calculated) Working volume: 15 - 175 µl
1.4	Material / Resin	PS (Polystyrene), free of heavy metal
1.5	Colour	Clear
1.6	Sterilization	No
1.7	Quality Control	- <u>Raw Material-Control</u> : physical and immunological testing - <u>Product-Control</u> : testing of attributive and variable characteristics in accordance with the valid specification
1.8	Other Information	For single use only

2.	Features	
2.1	Basic features	Free of detectable DNase/RNase, human DNA and pyrogens.
2.2	Temperature range	-20°C to +60°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	4800 x g: Swinging-bucket rotor
2.5	Chemical Resistance	See homepage: <a href="https://www.gbo.com/en_INT/know-how-services/download-center.html">https://www.gbo.com/en_INT/know-how-services/download-center.html</a>
2.6	Shelf life	4 years after month of production
2.7	Other Information	-

3.	Packaging	
3.1	Pieces / Bag	10
3.2	Pieces / Box	40
3.3	Lot-No.	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)
3.4	Other Information	Certificate of Quality

4.	Other Information
	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	<b>CONFIDENTIAL:</b> Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 02	Date 1 December 2014	Date 2 December 2014	Date 2 December 2014	
Date 14.12.2009	Name S. Kaelberer	Name Dr. R. Heller	Name 2 December 2014	

**DISCLAIMER:** The description of a certain product can only be considered as a guidance, because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.