1.	**Description / Specification**	
1.1	Description	PS Microplate, 96 well, half area well profile, solid bottom, alphanumeric well coding *675101*: standard style *675161*: standard style, sterile *675180*: physical surface treatment, with standard lid, sterile
1.2	Dimensions	See customer drawing
1.3	Volume	Total volume: 199 µl (mathematically calculated) Working volume: 15 - 175 µl *675180*: growth area: 0,15 cm^2
1.4	Material / Resin	Plate: PS (Polystyrene), free of heavy metal Lid: PS (Polystyrene), free of heavy metal
1.5	Colour	Plate: clear Lid: clear
1.6	Sterilization	*675101*: no *675161, -180*: SAL 10^{-3}
1.7	Quality Control	- Raw Material-Control: physical testing - Product-Control: testing of attributive and variable characteristics in accordance with the valid specification
1.8	Other Information	For single use only

2.	**Features**	
2.1	Basic features	All Item-No.: free of detectable DNase/RNase, human DNA and pyrogens. *675180*: contents non-cytotoxic
2.2	Temperature range	-20°C to +60°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	4800 x g: swinging-bucket rotor
2.5	Chemical Resistance	See homepage: https://www.gbo.com/en_INT/know-how-services/download-center.html
2.6	Shelf life	*675101*: - *675161*: 5 years after month of production *675180*: 4 years after month of production
2.7	Other Information	-

3.	**Packaging**	**675101, -161**	**675180**
3.1	Pieces / Bag	10	8
3.2	Pieces / Box	40	32
3.3	Lot-No.	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)	
3.4	Other Information	Certificate of Quality	

4.	**Other Information**
	-

Data Sheet subject to change without notice!

Prior Issue	**Drawn**	**Approved**	**Released**	**CONFIDENTIAL:** Information contained in this document or drawing is confidential and proprietory to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 04	Date 1 December 2014	Date 2 December 2014	Date 2 December 2014	
Date 14.12.2009	Name S. Kaelberer	Name Dr. R. Heller	Name A. Schulz	