PDS No. 6789xx	PRODUCT DATA SHEET				Page 1 of 1
Revision 05	CELLdisc™, Advanced TC™, Sterile				6
	Greiner Item-No. 6789xx				greiner bio-one
Valid for Item-No.:	678904 (sterile)	678908 (sterile)	678916 (sterile)	678940 (sterile	

1.	Description / Specification	
1.1	Description	Celldisc [™] multilayer device, from 1.000 cm ² to 10.000 cm ² , Advanced TC [™] treated, ventilation through central gas support channel, gas inlet and outlet equipped with filters, blue screw cap, sterile 678904: 4 layers, 1.000 cm ² 678908: 8 layers, 2.000 cm ² 678916: 16 layers, 4.000 cm ² 678940: 40 layers, 10.000 cm ²
1.2	Dimensions	See Customer Drawings
1.3	Volume	Working volume: 678904: 60-280 ml 678908: 120-560 ml 678916: 240-1120 ml 678940: 600-2800 ml <u>Growth area</u> : 678904: 678904: 1.000 cm² 678908: 2.000 cm² 678916: 4.000 cm² 678940: 10.000 cm²
1.4	Material / Resin	Disc : PS (Polystyrene), free of heavy metal Cap: HDPE (High Density Polyethylene), free of heavy metal Filter: PE (Polyethylene, free of heavy metal
1.5	Colour	Disc: clear Cap: blue Filter: nature
1.6	Sterilization	SAL 10 ⁻⁶
1.7	Quality Control	 <u>Raw Material-Control</u>: physical testing <u>Product-Control</u>: testing of attributive and variable characteristics in accordance with the valid specification
1.8	Other Information	For single use only

2.	Features	
2.1	Basic features	Free of detectable DNase/RNase, human DNA and pyrogens. Contents non-cytotoxic
2.2	Temperature range	4°C to +37 °C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	-
2.5	Chemical Resistance	See homepage: https://www.gbo.com/en_INT/know-how-services/download-center.html
2.6	Shelf life	2 years after month of production
2.7	Other Information	USP Class VI tested

3.	Packaging	678904	678908	678916	678940
3.1	Pieces / Bag	1	1	1	1
3.2	Pieces / Box	4	3	2	1
3.3	Lot-No.	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)			
3.4	Other Information	Certificate of Quality			

4.	Other Information	
4.1	For Laboratory use only	

Data Sheet subject to change without notice!						
Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this		
Revision 04	Date 17 January 2018	Date 17 January 2018	Date 17 January 2018	document or drawing is confidential and proprietory to Greiner Bio-One GmbH. This document may not be reproduced for any		
Date 18.07.2017	Name N. Kerzinger	Name Dr. L. Breth	Name A. Schulz	reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.		

Data Sheet subject to change without notice!

DISCLAIMER: The description of a certain product can only be considered as a guidance, because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.