


PDS No. 6800xx	PRODUCT DATA SHEET				Page 1 of 1
Revision 04	Roller Bottle CELLMASTER™, 1 X, PS				 greiner bio-one
	Greiner Item-No. 6800xx				
Valid for Item-No.:	680045 (sterile)	680048 (sterile)	680058 (sterile)	680060 (sterile)	680065 (sterile)
	680068 (sterile)				

1.	Description / Specification	
1.1	Description	Roller Bottle, short, smooth surface, graduation (scale 100-2000 ml), physical surface treatment, sterile <i>680048, -060, -065</i> : standard screw cap <i>680045, -058, -068</i> : filter screw cap
1.2	Dimensions	See customer drawing
1.3	Volume	Max. volume: 2520 ml Growth area: 850 cm ²
1.4	Material / Resin	Bottle: PS (Polystyrene), free of heavy metal <u>Screw cap</u> : HDPE (High Density Polyethylene), free of heavy metal <u>Filter</u> : PET (Polyethylene Terephthalat), PTFE (Polytetrafluorethane), free of heavy metal
1.5	Colour	<u>Bottle</u> : clear; print of graduation: black; writing area: white <u>Standard and filter screw cap</u> : blue <u>Filter</u> : white
1.6	Sterilisation	SAL 10 ⁻³
1.7	Quality Control	- <u>Raw Material-Control</u> : physical testing - <u>Product-Control</u> : testing of attributive and variable characteristics in accordance with the valid specification
1.8	Other Information	- For single use only - Expiry date and Lot-No. printed on label of bottle

2.	Features	
2.1	Basic features	Free of detectable DNase/RNase, human DNA and pyrogens. Contents non-cytotoxic. <i>680045, -058, -068</i> : controlled gas exchange through filter cap
2.2	Temperature range	-20°C to +60 °C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	N/A
2.5	Chemical Resistance	See homepage: www.gbo.com/bioscience →Products →Literature →Technical Information →Chemical Resistance of Resins
2.6	Shelf life	4 years after month of production
2.7	Other Information	Bottle and cap: USP Class VI tested

3.	Packaging	680045, -048	680058, -060	680065, -068
3.1	Pieces / Bag	24	2	24
3.2	Pieces / Box	48	24	24
3.3	Lot-No.	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)		
3.4	Other Information	Certificate of Quality		

4.	Other Information
	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 03	Date 21 October 2013	Date 21 October 2013	Date 21 October 2013	
Date 15.12.2009	Name S. Kaelberer	Name Dr. T. Schreiber	Name A. Schulz	