


PDS No.7101xx	PRODUCT DATA SHEET			Page 1 of 1
Revision 07	Pipettes, CELLSTAR [®] , 2 ml			 greiner bio-one
	Greiner Item-No. 7101xx			
Valid for Item-No.:	710107 (sterile)	710160 (sterile)	710180 (sterile)	710183 (sterile)

1.	Description / Specification	
1.1	Description	PS Pipette, 2 ml, sterile 710107, -160, -180: <u>Serological Pipettes</u> : graduation (scale 1/100 ml), green code ring, cotton plug 710183: <u>Aspiration Pipette</u> : standard (without plug, graduation, code ring)
1.2	Dimensions	See customer drawing
1.3	Volume	2 ml; volumetric accuracy +/- 2 %
1.4	Material / Resin	<u>Pipette</u> : PS (Polystyrene), free of heavy metal <u>Plug</u> : cotton roving
1.5	Colour	<u>Pipette</u> : clear <u>Plug</u> : white
1.6	Sterilization	SAL 10 ⁻³
1.7	Quality Control	<u>Product-Control</u> : testing of attributive and variable characteristics in accordance with the valid specification
1.8	Other Information	For single use only

2.	Features	
2.1	Basic features	Free of detectable DNase/RNase, human DNA and pyrogens. Contents non-cytotoxic.
2.2	Temperature range	0°C to +40°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	N/A
2.5	Chemical Resistance	See homepage: https://www.gbo.com/en_INT/know-how-services/download-center.html
2.6	Shelf life	5 years after month of production
2.7	Other Information	-

3.	Packaging	710107	710160	710180, -183
3.1	Pieces / Bag	25 Bulk packs (plastic)	1 / 100 "peel off"-packaging (plastic/plastic)	1 / 100 "peel off"-packaging (paper/plastic)
3.2	Pieces / Box	1000	1000	1000
3.3	Lot-No.	123456 (1=yr, 2&3=week number, 4=machine number, 5&6=pallet number)		
3.4	Other Information	Certificate of Quality		

4.	Other Information
	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 06	Date 13 October 2014	Date 14 October 2014	Date 15 October 2014	
Date 09.09.2013	Name S. Kaelberer	Name Dr. S. Mühlfriedel	Name A. Schulz	

DISCLAIMER: The description of a certain product can only be considered as a guidance, because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.