


PDS No. 76207x	PRODUCT DATA SHEET				Page 1 of 1
Revision 06	96 Well Strip Plate, PS, F-Bottom				 greiner bio-one
	Greiner Item-No. 76207x				
Valid for Item-No.:	762070	762071	762074	762075	762076
	762077				

1.	Description / Specification	
1.1	Description	PS Strip plate, 12 x F8 strips mounted in frame, solid F-bottom (flat) 762070: MICROLON [®] 200, medium binding 762071: MICROLON [®] 600, high binding 762074: LUMITRAC [™] 600, high binding 762075: LUMITRAC [™] 200, medium binding 762076: FLUOTRAC [™] 200, medium binding 762077: FLUOTRAC [™] 600, high binding
1.2	Dimensions	See customer drawing
1.3	Volume per well	Total volume: 388 µl Working volume: 20-350 µl
1.4	Material / Resin	Strips and Frame: PS (Polystyrene), free of heavy metal
1.5	Colour	Frame: white 762070, -071: strip: clear 762074, -075: strip: white 762076, -077: strip: black
1.6	Sterilization	No
1.7	Quality Control	- Raw Material-Control: physical and immunological testing - Product-Control: testing of attributive and variable characteristics in accordance with the valid specification
1.8	Other Information	For single use only

2.	Features	
2.1	Basic features	Free of detectable DNase/RNase, human DNA and pyrogens.
2.2	Temperature range	-20°C to +60°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	N/A
2.5	Chemical Resistance	See homepage: https://www.gbo.com/en_INT/know-how-services/download-center.html
2.6	Shelf life	4 years after month of production
2.7	Other Information	-

3.	Packaging	
3.1	Pieces / Bag	10
3.2	Pieces / Box	100
3.3	Lot-No.	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)
3.4	Other Information	Certificate of Quality

4.	Other Information
	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 05	Date 2 December 2014	Date 3 December 2014	Date 3 December 2014	
Date 15.07.2014	Name S. Kaelberer	Name Dr. R. Heller	Name A. Schulz	

DISCLAIMER: The description of a certain product can only be considered as a guidance, because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.