


PDS No. 76707x	PRODUCT DATA SHEET			Page 1 of 1
Revision 04	96 Well ELISA-Plate, PS, 12 x U8 Strips, U-Bottom			
	Greiner Item-No. 76707x			
Valid for Item-No.:	767070	767071		

1.	Description / Specification	
1.1	Description	PS Strip plate ELISA, 12 x U8 strips mounted in frame, solid U-bottom (round), alphanumeric well coding 767070: MICROLON [®] 200, medium binding 767071: MICROLON [®] 600, high binding
1.2	Dimensions	See customer drawing
1.3	Volume per well	Total volume: 312 µl Working volume: 50 - 280 µl
1.4	Material / Resin	Strips and Frame: PS (Polystyrene), free of heavy metal
1.5	Colour	Frame: white Strips: clear
1.6	Sterilization	No
1.7	Quality Control	- <u>Raw Material-Control</u> : physical and immunological testing - <u>Product-Control</u> : testing of attributive and variable characteristics in accordance with the valid specification
1.8	Other Information	For single use only

2.	Features	
2.1	Basic features	Free of detectable DNase/RNase, human DNA and pyrogens.
2.2	Temperature range	-20°C to +60°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	N/A
2.5	Chemical Resistance	See homepage: https://www.gbo.com/en_INT/know-how-services/download-center.html
2.6	Shelf life	4 years after month of production
2.7	Other Information	-

3.	Packaging	
3.1	Pieces / Bag	10
3.2	Pieces / Box	100
3.3	Lot-No.	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)
3.4	Other Information	Certificate of Quality

4.	Other Information
	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 03	Date 2 December 2014	Date 3 December 2014	Date 3 December 2014	
Date 15.07.2014	Name S. Kaelberer	Name Dr. U. Honisch	Name A. Schulz	

DISCLAIMER: The description of a certain product can only be considered as a guidance, because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.