


PDS No. 77235x	PRODUCT DATA SHEET				Page 1 of 1
Revision 00	Sapphire Pipette Tips, 10 µl, XL, Universal				
	Greiner Item-No. 77235x				
Valid for Item-No.:	772350	772351	772352 (sterile)	772354	772355 (sterile)

1.	Description / Specification	
1.1	Description	PP Universal Sapphire Pipette Tip, 10 µl, XL, graduated For liquid handling of volumes in the range of 0.5 µl up to 10 µl, universal fit, graduated, extended tip for better recovery of small sample volumes 772350: bag, non-sterile 772351: rack, non-sterile 772352: rack, sterile 772354: refill, non-sterile 772355: refill, sterile Matching rack: Cat.-No. 770310 (red insert)
1.2	Dimensions	See Customer Drawing
1.3	Volume	0.5-10 µl
1.4	Material / Resin	Pipette tip: PP (Polypropylene), free of heavy metal Rack/lid: PP (Polypropylene), free of heavy metal
1.5	Colour	Pipette tip: translucent Rack/lid: translucent; Rack insert: red
1.6	Sterilization	772350, -351, -354: No 772352, -355: by irradiation [SAL10 ⁶]
1.7	Quality Control	- <u>Raw Material control</u> : physical testing - <u>Product control</u> : testing of attributive and variable characteristics in accordance with the valid specification
1.8	Other Information	- For single use only - Universal tip suitable for commonly used laboratory pipettors such as Eppendorf, Gilson, Sapphire, Brand - see valid compatibility list

2.	Features	
2.1	Basic features	Free of detectable DNase/RNase, human DNA, PCR inhibitors and endotoxins (pyrogens)
2.2	Temperature range	Room Temperature (15°C - 30°C)
2.3	Autoclavability	772350, -351, -354: 121°C / 20 min. 772352, -355: n.a. (sterile by irradiation)
2.4	Centrifugation, max. RCF	N/A
2.5	Chemical Resistance	See homepage: https://www.gbo.com/en_INT/know-how-sevices/download-center.html
2.6	Shelf life	3 years
2.7	Other Information	Compatibility testing procedure of the pipette tips follows the ISO 8655

3.	Packaging	772350	772351, 772352	772354, 772355
3.1	Pieces / packaging unit	1000 (bag)	96 (rack)	96 (rack / refill unit)
3.2	Pieces / Box	-	960	960 (refill unit)
3.3	Pieces / Case	5.000	5.760	4.800
3.4	Lot-No.	A YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)		
3.5	Other Information	Certificate of Quality to download		

4.	Other Information
-	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision	Date	Date	Date	
-	4 November 2020	16 November 2020	20 November 2020	
Date	Name	Name	Name	
-	S. Kaelberer	J. Gaiser	A. Illig	

DISCLAIMER: The description of a certain product can only be considered as a guidance, because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.