


PDS No. 772363	<b>PRODUCT DATA SHEET</b>		Page 1 of 1
Revision 00	Sapphire Filter Pipette Tips, 10 µl, XL, Universal, Low Retention		
	Greiner Item-No. 772363		
Valid for Item-No.:	<b>772363</b> (sterile)		

1.	Description / Specification	
1.1	Description	PP Universal, Low Retention, Sapphire Filter Pipette Tip 10 µl, XL, graduated For liquid handling of volumes in the range of 0.5 µl up to 10 µl, universal fit, graduated, extended tip for better recovery of small sample volumes, with rack, sterile
1.2	Dimensions	See Customer Drawing
1.3	Volume	0.5-10 µl
1.4	Material / Resin	Pipette tip: PP (low retention Polypropylene), free of heavy metal Filter: PE (Polyethylene), average pore size: 16-35 µm Rack/lid: PP (Polypropylene), free of heavy metal
1.5	Colour	Pipette tip: translucent Filter: white Rack/lid: translucent; Rack insert: red
1.6	Sterilization	By irradiation [SAL 10 <sup>-6</sup> ]
1.7	Quality Control	- Raw Material control: physical testing - Product control: testing of attributive and variable characteristics in accordance with the valid specification
1.8	Other Information	- For single use only - Universal tip suitable for commonly used laboratory pipettors such as Eppendorf, Gilson, Sapphire, Brand - see valid compatibility list

2.	Features	
2.1	Basic features	Free of detectable DNase/RNase, human DNA, PCR inhibitors and endotoxins (pyrogens)
2.2	Temperature range	Room Temperature (15°C - 30°C)
2.3	Autoclavability	N/A (sterile by irradiation)
2.4	Centrifugation, max. RCF	N/A
2.5	Chemical Resistance	See homepage: <a href="https://www.gbo.com/en_INT/know-how-sevices/download-center.html">https://www.gbo.com/en_INT/know-how-sevices/download-center.html</a>
2.6	Shelf life	3 years
2.7	Other Information	Compatibility testing procedure of the pipette tips follows the ISO 8655

3.	Packaging	772363		
3.1	Pieces / packaging unit	96 (rack)		
3.2	Pieces / Box	960		
3.3	Pieces / Case	5.760		
3.3	Lot-No.	A YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)		
3.4	Other Information	Certificate of Quality to download		

4.	Other Information
	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision	Date	Date	Date	
-	3 November 2020	16 November 2020	19 November 2020	
Date	Name	Name	Name	
-	S. Kaelberer	J. Gaiser	A. Illig	

DISCLAIMER: The description of a certain product can only be considered as a guidance, because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.