PDS No. 77635x	PRODUCT DATA SHEET				Page 1 of 1
Dovision 01	Sapphire Pipette Tips, 300 µl, Universal				0
Revision 01	Greiner Item-No. 77635x				greiner
Valid for Item-No.:	776350	776351	776352 (sterile)	776354	776355 (sterile)

1.	Description / Specification	
For liquid handling of volumes in the range of 10 µ graduated 776350: bag, non-sterile		776350: bag, non-sterile
		776351: rack, non-sterile 776352: rack, sterile 776354: refill, non-sterile 776355: refill, sterile
		Matching Rack: 770330 (green insert)
1.2	Dimensions	See Customer Drawing
1.3	Volume	10-300 µl
1.4	Material / Resin	Pipette tip: PP (Polypropylene), free of heavy metal Rack/lid: PP (Polypropylene), free of heavy metal
1.5	Colour	Pipette tip: translucent Rack/lid: translucent; Rack insert: green
1.6	Sterilization	776350, -351, -354: No 776352, -355: by irradiation [SAL10 ⁻⁶]
1.7	Quality Control	Raw Material control: physical testing Product control: testing of attributive and variable characteristics in accordance with the valid specification
1.8	Intended Use	General laboratory product for pipetting of liquids to be used by qualified personnel in a laboratory environment.
1.9	Other Information	- For single use only - Universal tip suitable for commonly used laboratory pipettors such as Eppendorf, Gilson, Sapphire, Brand - see valid compatibility list

2.	Features		
2.1	Basic features	Free of detectable DNase/RNase, human DNA, PCR inhibitors and endotoxins (pyrogens)	
2.2	Temperature range	Room Temperature (15°C - 30°C)	
2.3	Autoclavability	776350, -351, -354: 121°C / 20 min. 776352, -355: n.a. (sterile by irradiation)	
2.4	Centrifugation, max. RCF	N/A	
2.5	Chemical Resistance	See homepage: https://www.gbo.com/en_INT/know-how-sevices/download-center.html	
2.6	Shelf life	3 years	
2.7	Other Information	Compatibility testing procedure of the pipette tips follows the ISO 8655	

3.	Packaging	776350	776351, 776352	776354, 776355
3.1	Pieces / packaging unit	1000 (bag)	96 (rack)	96 (rack / refill unit)
3.2	Pieces / Box	-	960	n.a.
3.3	Pieces / Case	10.000	5.760	4.800
3.3	Lot-No.	A YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)		
3.4	Other Information	Certificate of Quality to download		

4 Ot	ther Information
- 01	

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this
Revision 00	Date 17 May 2021	Date 18 May 2021	Date 19 May 2021	document or drawing is confidential and proprietory to Greiner Bio-One GmbH. This document may not be reproduced for any
Date 20.11.2020	Name S. Kaelberer	Name J. Gaiser	Name A. Illig	reason without written permission from Greine Bio-One GmbH. All rights of design, invention and copyright are reserved.