

Specification

Solid medium for the detection of urease, according to ISO standards and DIN standard.

Presentation

20 Tubes / Slant
Tube 16 x 113 mm
with: 6,2 ± 0,3 ml

Packaging Details

1 box with 20 tubes, 16x113 mm glass tubes, ink
labelled and metal cap.

Shelf Life

9 months

Storage

8-14°C

Composition

Composition (g/l):

Peptone.....	1.000
D(+) Glucose.....	1.000
Sodium chloride.....	5.000
Disodium phosphate.....	1.200
Potassium dihydrogen phosphate.....	0.800
Phenol red.....	0.012
Urea 40%.....	50 ml
Agar.....	15.000

Description /Technique

Urea Agar complies with Christensen's specifications, and is recommended for the detection of ureolytic or urea degrading microorganisms, especially Enterobacteriaceae, although it can be used with Gram positive bacteria.

A pure culture is inoculated by surface streaking, and then incubated at 35±2°C. Generally, organisms with strong urease activity can be read after 4-6 hours. Reaction is evident as the medium changes colour from orange to pink-fuchsia, due to a strong alkalization produced by ammonia release.

Quality control

Physical/Chemical control

Color : Orange pH: 6.8 ± 0.2 at 25°C

Microbiological control

Loop spreading

Aerobiosis. Incubation at 35 ± 2°C, reading at 6-18 h

Microorganism

Escherichia coli ATCC® 25922, WDCM 00013
Salmonella typhimurium ATCC® 14028, WDCM 00031
Proteus mirabilis ATCC® 43071
Proteus hauseri ATCC® 13315
Enterobacter aerogenes ATCC® 13048, WDCM 00175

Growth

Good - Ureasa negative
Good - Negative urea
Good - Ureasa positive
Good - Ureasa positive
Good - Ureasa negative

Sterility Control

Incubation 48 hours at 30-35°C and 48 hours at 20-25°C: NO GROWTH
Check at 7 days after incubation in same conditions

Bibliography

- ATLAS, R.M. & L.C. PARK (1993) Handbook of Microbiological Media. CRC Press Inc. London.
- CHRISTENSEN W.B. (1946) Urea decomposition as means of differentiating *Proteus* and *Paracolon* cultures from each other and from *Salmonella* and *Shigella* types. J. Bact. 52:461.
- DIN Standard 10160. Untersuchung von Fleisch und Fleischerzeugnissen. Nachweis von Salmonellen. Referenzverfahren.
- DOWNES, F.P. & K. ITO (2001) Compendium of methods for the microbiological examination of foods. 4th ed. APHA. Washington DC. USA.
- EDWARDS & EWING (1962) Identificacion of Enterobacteriaceae Burgess Pub. Co.
- FIL-IDF 93 Standard (2001) Milk and Milk products. Detection of *Salmonella*.
- ISO 6340 Standard (1995) Water Quality - Detection of *Salmonella* spp.
- ISO Standard 6579-1 (2017) Microbiology of food chain - Horizontal method for the detection, enumeration and serotyping of *Salmonella* - Part 1 : Detection of *Salmonella* spp.
- ISO 6785 Standard (2001) Milk and Milk products - Detection of *Salmonella* spp.
- ISO 21567 Standard (2004) Microbiology of food and animal feeding stuffs.- Horizontal method for the detection of *Shigella* spp.
- MARSHALL, R.T. (1992) Standard methods for the examination of dairy products. 16th ed. APHA. Washington DC. USA.
- ISO 11133:2014/ Adm 1:2018. Microbiology of food, animal feed and water. Preparation, production, storage and performance testing of culture media.
- UNE-EN ISO 11133 (2014). Microbiología de los alimentos para consumo humano, alimentación animal y agua.-Preparación, producción, conservación y ensayos de rendimiento de los medios de cultivo.