

# Phosphate

**Test kit for performing colorimetric tests  
on phosphate ions in surface water and sewage**

**Method:**

Ammonium molybdate forms with phosphate ions phosphomolybdic acid, which is reduced to phosphomolybdenum blue.

**Measurement range:**

0.2–5 mg/L PO<sub>4</sub>-P

**Contents of test kit (\*refill pack):**

sufficient for 80 tests

- 25 mL PO<sub>4</sub>-1\*
- 25 mL PO<sub>4</sub>-2\*
- 2 screw-plug measuring glasses
- 1 slide comparator
- 1 colour chart
- 1 plastic syringe 5 mL
- 1 instructions for use\*

**Hazard warning:**

Information regarding safety can be found on the box' label and in the safety data sheet. You can download the SDS from [www.mn-net.com/SDS](http://www.mn-net.com/SDS).

**Instructions for use:****a) colorimetric determination with color chart**

also refer to the pictogram on the back of the color chart

1. Pour a **5 mL water sample** into each of the measuring glasses using the plastic syringe.  
Place a measuring glass on position A in the comparator.

**Only add the reagent to measuring glass B.**

2. Add **6 drops of PO<sub>4</sub>-1**, seal the glass and mix.
3. Add **6 drops of PO<sub>4</sub>-2**, seal the glass and mix.
4. Open the glass after **10 min** and place it on position B in the comparator.
5. Slide the comparator until the colors match in the inspection hole on top. Check the measurement reading in the recess on the comparator reed. Mid-values can be estimated.
6. After use, rinse out both measuring glasses thoroughly and seal them.

**b) photometric determination**

The reagents are also suitable for **photometric evaluation**. Please refer to the separate instructions for photometric performance.

This technique can be used also for analyzing sea water.

**Disposing of the samples:**

Information regarding disposal can be found in the safety data sheet. You can download the SDS from [www.mn-net.com/SDS](http://www.mn-net.com/SDS).

**Interferences:**

Larger amounts of oxidizing reagents inhibit formation of the blue color complex and have to be destroyed. H<sub>2</sub>S interferes in concentrations above 2 mg/L, but can be expelled after acidification of the water sample. Heavy metals in excess of 10 mg/L can slightly decrease the intensity of the color (vanadium causes an increase in color). Silica interferes in excess of 10 mg/L Si.

**Conversion table:**

mg/L PO <sub>4</sub> -P	mg/L PO <sub>4</sub> <sup>3-</sup>	mg/L P <sub>2</sub> O <sub>5</sub>
0.2	0.6	0.5
0.3	0.9	0.7
0.5	1.5	1.1
0.7	2.1	1.6
1	3	2
2	6	5
3	9	7
5	15	12

**Storage:**

Store the test kit in a cool (< 25 °C) and dry place.