PDS No. 78189x	S No. 78189x PRODUCT DATA SHEET		
Revision 06	384 Well SensoPlate™, PS, Glass Bottom	6	
	Greiner Item-No. 78189x	greiner bio-one	
Valid for Item-No.:	781892 (sterile) 781896 (sterile)		

1.	Description / Specification			
1.1	Description	PS Microplate, 384 well, F-glass bottom (flat), rounded square well design, alphanumeric well coding, sterile 781892: with single position lid, low profile		
1.2	Dimensions	781896: without lid		
1.2	Dimensions	See customer drawing Glass bottom: 175 μm (+/- 20 μm)		
1.3	Volume per well	Total volume: 138 μl (mathematically calculated) Working volume: 10 – 130 μl		
1.4	Material / Resin	Plate: PS (Polystyrene), free of heavy metal Glass bottom: clear borosilicate Lid: PS (Polystyrene), free of heavy metal		
1.5	Colour	Plate: black Glass bottom: clear Lid: clear		
1.6	Sterilization	SAL 10 ⁻³		
1.7	Quality Control	 <u>Raw Material-Control</u>: physical testing <u>Product-Control</u>: testing of attributive and variable characteristics in accordance with the valid specification 		
1.8	Other Information	For single use only		

2.	Features	
2.1	Basic features	Adhesive: compatible with cell culture media, low autofluorescence
2.2	Temperature range	+ 4°C to +37°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	N/A
2.5	Shelf life	2 years after month of production
2.6	Other Information	-

3.	Packaging	
3.1	Pieces / Air Cushion Bag	1
3.2	Pieces / Box	16
3.3	Lot-No.	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)
3.4	Other Information	Certificate of Quality

4.	Other Information	
	-	

		Data Sheet subject to cha	inge without notice!		
Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this	
Revision	Date	Date	Date	document or drawing is confidential and proprietory to Greiner Bio-One GmbH. This	
05	1 February 2016	2 February 2016	3 February 2016	document may not be reproduced for any	
Date	Name	Name	Name	reason without written permission from Greiner Bio-One GmbH. All rights of design, invention,	
03.12.2014	S. Kaelberer	Dr. R. Heller	Dr. A. Ganser	and copyright are reserved.	

DISCLAIMER: The description of a certain product can only be considered as a guidance, because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.