


PDS No. 7852xx	PRODUCT DATA SHEET			Page 1 of 1
Revision 06	Sapphire PCR Microplates, PP, 384 Well, Full Skirt			
	Greiner Item-No. 7852xx			
Valid for Item-No.:	785201	785290		

1.	Description / Specification	
1.1	Description	Sapphire PCR Microplate, 384 well with full skirt, alphanumeric well coding 785201: standard 785290: suitable for ABI Suitable for adhesive sealers and heat-sealing
1.2	Dimensions	See Customer Drawings
1.3	Volume per well	Total volume: 45 µl Working volume: 25 µl
1.4	Material / Resin	PP (Polypropylene), free of heavy metal
1.5	Colour	Translucent, blue alphanumeric well coding
1.6	Sterilization	No
1.7	Quality Control	<u>Raw Material-Control</u> : physical testing <u>Product-Control</u> : testing of attributive and variable characteristics in accordance with the valid specification
1.8	Other Information	For single use only

2.	Features	
2.1	Basic features	Free of detectable DNase/RNase, human DNA and pyrogens
2.2	Temperature range	-80°C to +105°C
2.3	Autoclavability	Not recommended
2.4	Centrifugation, max. RCF	4800 x g: swinging-bucket rotor
2.5	Chemical Resistance	See homepage: https://www.gbo.com/en_INT/know-how-services/download-center.html
2.6	Shelf life	N/A
2.7	Other Information	-

3.	Packaging	
3.1	Pieces / Box	15
3.2	Pieces / Case	60
3.3	Lot-No.	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)
3.4	Other Information	Certificate of Quality

4.	Other Information
	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 05	Date 13 April 2016	Date 04 May 2016	Date 04 May 2016	
Date 04.12.2014	Name S. Kaelberer	Name Dr. R. Heller	Name Dr. A Ganser	

DISCLAIMER: The description of a certain product can only be considered as a guidance, because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.