EC Declaration of Conformity

WIR / WE / NOU

DITABIS

DIGITAL BIOMEDICAL IMAGING SYSTEM AG

Anschrift / Address / Adresse

Freiburger Str. 3
D – 75179 Pforzheim

declare on our own authority that the following product(s)

Thermoshaker

MKR 13, MKR 23, MHR 13, MHR 23, MHL 23, MHR 11

BlockThermostats

TK 23, TH 21

to which this declaration relates is following the provisions of the following standards or normative documents:

IEC 61010-1:2010+Cor.2011
Safety requirements for electrical equipment for measurement, control and laboratory use - Part 1: General requirements

IEC 61010-2-010:2003
Safety requirements for electrical equipment for measurement, control and laboratory use - Part 2-010: Particular requirements for laboratory equipment for the heating of materials

IEC 61326-1:2005
Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 1: General requirements

in accordance with provisions of the Directives

2004/108/EC (EMC Guideline)
2006/95/EC (EU Low Voltage Guideline)
2011/65/EC (Restriction of use of hazardous substances)

This Declaration of Conformity relates to all product units which are supplied from the release date stated below.

Pforzheim, 28-05-2013

.....
Aaron Taylor (CEO)