


PDS No.8032xx	PRODUCT DATA SHEET			Page 1 of 1
Revision 02	Datamatrix-Cryo-Rack, PP			
	Greiner Item-No. 8032xx			
Valid for Item-No.:	803202	803270	803277	

1.	Description / Specification	
1.1	Description	Storage Cryo Rack with a footprint that corresponds to the American National Standards Institute (ANSI-Standard), capacity for 48 Cryo.s™, stackable, rack bottom with scanning windows for the Datamatrix of the contained Cryo.s™ 803202: rack with low lid (suitable for 1 ml and 2 ml Cryo.s™) 803270: rack with high lid (suitable for 4 ml and 5 ml Cryo.s™) 803277: rack without lid (suitable for 1 ml, 2 ml, 4 ml and 5 ml Cryo.s™)
1.2	Dimensions	See customer drawing
1.3	Volume	N/A
1.4	Material / Resin	PP (Polypropylene), free of heavy metal
1.5	Colour	Rack: black Lids: translucent
1.6	Sterilization	No
1.7	Quality Control	Product-Control: testing of attributive and variable characteristics in accordance with the valid specification
1.8	Other Information	-

2.	Features	
2.1	Basic features	-
2.2	Temperature range	-196°C to ~ +30°C
2.3	Autoclavability	N/A
2.4	Centrifugation, max. RCF	N/A
2.5	Chemical Resistance	See homepage: https://www.gbo.com/en_INT/know-how-services/download-center.html
2.6	Shelf life	N/A
2.7	Other Information	-

3.	Packaging	803202	803270	803277
3.1	Pieces / Bag	5	15	5
3.2	Pieces / Box	20	15	20
3.3	Lot-No.	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)		
3.4	Other Information	-		

4.	Other Information
	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 01	Date 3 December 2014	Date 4 December 2014	Date 4 December 2014	
Date 11.11.2010	Name S. Kaelberer	Name Dr. S. Mühlfriedel	Name A. Schulz	

DISCLAIMER: The description of a certain product can only be considered as a guidance, because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.