

4X CAPITAL™ 1-Step qRT-PCR Probe Master Mix

LOT: See product label

EXPIRY DATE: See product label

ORDERING INFORMATION

CAT. NO.	SIZE	PACKAGE CONTENT
BR0502001	200 reactions of 20 µl	1 ml CAPITAL qPCR Probe Master Mix (1step) 200 µl RTase with RNase Inhibitor
BR0502002	1000 reactions of 20 µl	5 × 1 ml CAPITAL qPCR Probe Master Mix (1step) 5 × 200 µl RTase with RNase Inhibitor
BR0502101	200 reactions of 20 µl	1 ml CAPITAL qPCR Probe Master Mix LROX (1step) 200 µl RTase with RNase Inhibitor
BR0502102	1000 reactions of 20 µl	5 × 1 ml CAPITAL qPCR Probe Master Mix LROX (1step) 5 × 200 µl RTase with RNase Inhibitor
BR0502201	200 reactions of 20 µl	1 ml CAPITAL qPCR Probe Master Mix HROX (1step) 200 µl RTase with RNase Inhibitor
BR0502202	1000 reactions of 20 µl	5 × 1 ml CAPITAL qPCR Probe Master Mix HROX (1step) 5 × 200 µl RTase with RNase Inhibitor

COMPONENT	COMPOSITION
4X CAPITAL 1-Step qPCR Probe Master Mix	Optimized 4X qPCR Probe Master Mix for One Step qRT-PCR
LROX Mix / HROX Mix	Rox incorporated in the mix in low / high concentration
RTase with RNase Inhibitor	Proprietary 20X Reverse transcriptase in a mix with efficient Ribonuclease Inhibitor

STORAGE

-30°C to -10°C (until expiry date – see product label)
Protect from light. Avoid multiple freeze thaw cycles by preparing aliquots.

FEATURES

- Best in-class performance for both single and multiplex detection
- Convenient master mix for detection of low-copy pathogen targets
- High specificity and sensitivity across a wide range of sample sources

APPLICATIONS

- One step qRT-PCR from mRNA, total RNA and viral RNA targets
- For use with standard and fast qPCR platforms
- Single and multiplex qRT-PCR reactions

4X CAPITAL™ 1-Step qRT-PCR Probe Master Mix

DESCRIPTION

biotechrabbit™ 4X CAPITAL 1-Step qRT-PCR Probe Master Mix provides outstanding performance for real-time PCR quantification of RNA templates, including mRNA, total RNA and viral RNA from a wide range of targets. The master mix ensures high specificity and sensitivity in single and multiplex detection, making it the choice for extremely low-copy-number targets in pathogen detection.

4X CAPITAL 1-Step qRT-PCR Probe Master Mix uses proprietary reverse transcriptase technology and buffer chemistry for efficient cDNA synthesis and QPCR in a single tube. To enable the use of the kit on qPCR platforms with different reference dye concentration requirements, three kit formats are available: a one-step kit containing no ROX, as well as LROX and HROX versions containing ROX in the corresponding concentrations.

Info: Recommended annealing temperature is 2°C above primer T_m (use gradient PCR to optimize the annealing temperature).

ROX REFERENCE DYE

- See PCR cyclers instruction for recommended concentration of ROX passive reference dye

PROTOCOL

Notes

- For efficient amplification under fast cycling conditions use amplicon lengths between 80 bp and 200 bp.
- The shorter the amplicon length the faster the reaction can be cycled. Use maximum 400 bp amplicons.
- Primers should have a predicted melting temperature of around 60°C, using default Primer 3 settings (<http://frodo.wi.mit.edu/primer3/>).
- For TaqMan® probes choose probe close to 5' primer, avoid terminal guanosine residues.

Prevention of reaction contamination

RNase contamination is an exceptional concern when working with RNA. RNase A, providing most threat to RNA integrity, is a highly stable contaminant of any laboratory. To prevent RNA from degradation and to minimize possibility of contamination One Step RT-PCR; follow the guidelines below:

- Use separate clean areas for preparation of the samples and the reaction mixture.
- DEPC-treat all tubes and pipette tips or use certified nuclease-free labware with aerosol filters.
- Wear fresh gloves when handling RNA and all reagents.
- Always assess the integrity of RNA prior to RT-PCR in denaturing agarose gel electrophoresis.
- Use only water and reagents that are free of DNA, DNases and RNases.
- With every One Step RT-PCR setup, perform a contamination control reaction without template DNA.

4X CAPITAL 1-Step qRT-PCR Probe Master Mix

Basic Protocol

- Keep the master mix protected from light until you use it.
- Aliquot the master mix to minimize freeze-thaw cycles and light exposure.
- Thaw on ice and mix very well all reagents. Assemble and keep all reactions on ice.
- Use only high quality optically clear reaction plates and seals designed for fluorescence applications.
- Do not use corner wells or use a more robust seal.
- Reserve plate positions for positive (control DNA) and negative (water or buffer) controls.
- First pipette the primer mixture, then add the template and last the Master Mix.
- Before preparing mixes, calculate the volume needed according to the reaction number plus one extra.
- To have a better correlation, run the reactions in triplets.

COMPONENT	VOLUME	FINAL CONCENTRATION
Primer Mix (Reverse and Forward)	Variable	100- 400 nM
<i>Too high primer concentrations result in unspecific amplification and should be avoided.</i>		
Specific Probe	Variable	200 nM
Template RNA	Variable	0.01 pg to 1 µg
<i>Use 1 pg – 1 µg Total RNA, or >0.01 pg mRNA</i>		
4X CAPITAL qPCR Probe Master Mix (1step)	5 µl	1×
20X RTase with RNase Inhibitor	1 µl	1×
Nuclease free water	Variable	
Total volume	20 µl	
<ul style="list-style-type: none">• Gently mix the reactions without creating bubbles (do not vortex). Bubbles will interfere with fluorescence detection. Place the reaction into the PCR cyclor.		

CYCLING PROGRAM

STEP	TEMPERATURE	TIME	CYCLES
Reverse Transcription	50°C	10 min	1
Initial activation	95°C	3 min	1
Denaturation	95°C	10 s	40-45
Annealing/Extension*	(60-68°C)*	30 s	

* Recommendation is primer T_m +2°C or use gradient PCR to optimize the annealing temperature. Do not use annealing temperatures below 60°C. For melt analysis refer to instrument instructions.

4X CAPITAL™ 1-Step qRT-PCR Probe Master Mix

CERTIFICATE OF ANALYSIS

Quality Control

Functional assay

Mix tested functionally in qRT-PCR.

Quality confirmed by: Head of Quality Control

SAFETY INSTRUCTIONS

For safety instructions please see Safety Data Sheets (SDS)/Sicherheitshinweise finden Sie in den SDS unter: <http://www.biotechrabbit.com/support/documentation.html>.

USEFUL HINTS

- Visit Applications at www.biotechrabbit.com for more products and product selection guides.
- Most biotechrabbit products are available in custom formulations and bulk amounts.

CONTACT BIOTECHRABBIT

biotechrabbit GmbH

Volmerstr. 9a
12489 Berlin,
Germany

info@biotechrabbit.com
support@biotechrabbit.com
www.biotechrabbit.com

Phone: +49 30 555 7821-10
Fax: +49 30 555 7821-99



Legal Disclaimer and Product Use Limitation

Purchase of product does not include a license to perform any patented applications; therefore it is the sole responsibility of users to determine whether they may be required to engage a license agreement depending upon the particular application in which the product is used. This product was developed, manufactured, and sold for in vitro use only. It is not suitable for administration to humans or animals.

Trademarks: biotechrabbit™, CAPITAL™ (biotechrabbit GmbH).

valid from 10.02.2021