



Product Service

# Attestation of Compliance

No. E8N 12 10 48128 047

**Holder of Certificate:** Hangzhou Bioer Technology Co., Ltd.

1192BinAn Rd, Binjiang District  
310053 Hangzhou  
PEOPLE'S REPUBLIC OF CHINA

**Name of Object:** Laboratory Equipment  
Thermal cycler

**Model(s):** TC-E-96G, TC-E-3048U, TC-E-384G, TC-E-48D,  
TC-E-4I, TC-96/G/H(b)A, Arktik 0196, Arktik 0138,  
Arktik 0148, SWT-MXP-BLC-1, SWT-MXP-BLC-2,  
SWT-MXP-BLC-3, SWT-MXP-BLC-4, SWT-MXP-BLC-5,  
TC-96/G/H(b)B, TC-EA-96GA, TC-EA-3048UA,  
TC-EA-384GA, TC-EA-48DA, TC-EA-4IA

**Description of Object:**

Rated voltage:	100-240V~
Rated frequency:	50/60Hz
Rated input power:	600W
Protection class:	I

**Tested according to:** EN 61326-1:2006

This Attestation of Compliance is issued according to the Directive 2004/108/EC relating to electromagnetic compatibility on a voluntary basis. It confirms that the listed apparatus complies with all essential requirements of the EMC directive and applies only to the sample and its technical documentation submitted to TÜV SÜD Product Service GmbH for testing and certification. See also notes overleaf.

**Test report no.:** 4842012224800

*Jun Bao*



**Date,** 2012-10-31 ( Jun Bao )

**CE** After preparation of the necessary technical documentation as well as the EC conformity declaration the required CE marking can be affixed on the product. Other relevant directives have to be observed.

Page 1 of 1