ZERTIFIKAT ◆ CERTIFICATE ◆ 認証証書 ◆ СЕРТИФИКАТ ◆ CERTIFICADO ◆ CERTIFICAT

Product Service

Attestation of Compliance

No. E8N 12 10 48128 047

Holder of Certificate: **Hangzhou Bioer Technology Co., Ltd.**
1192BinAn Rd,Binjiang District
310053 Hangzhou
PEOPLE'S REPUBLIC OF CHINA

Name of Object: **Laboratory Equipment Thermal cycler**

Model(s): **TC-E-96G,TC-E-3048U,TC-E-384G,TC-E-48D, TC-E-4I,TC-96/G/H(b)A,Arktik 0196,Arktik 0138, Arktik 0148,SWT-MXP-BLC-1,SWT-MXP-BLC-2, SWT-MXP-BLC-3,SWT-MXP-BLC-4,SWT-MXP-BLC-5, TC-96/G/H(b)B,TC-EA-96GA,TC-EA-3048UA, TC-EA-384GA,TC-EA-48DA,TC-EA-4IA**

Description of Object:

Rated voltage:	100-240V~
Rated frequency:	50/60Hz
Rated input power:	600W
Protection class:	I

Tested according to: EN 61326-1:2006

This Attestation of Compliance is issued according to the Directive 2004/108/EC relating to electromagnetic compatibility on a voluntary basis. It confirms that the listed apparatus complies with all essential requirements of the EMC directive and applies only to the sample and its technical documentation submitted to TÜV SÜD Product Service GmbH for testing and certification. See also notes overleaf.

Test report no.: 4842012224800

Jun Bao

TÜV SÜD 545295

Date, 2012-10-31 (Jun Bao)

CE After preparation of the necessary technical documentation as well as the EC conformity declaration the required CE marking can be affixed on the product. Other relevant directives have to be observed.

TÜV®