

XI EMC



Product Service

EC-Attestation of Compliance

No. E8N 09 05 48128 026

Holder of Certificate: Hangzhou Bioer Technology Co., Ltd.

1192BinAn Rd, Binjiang District
310053 Hangzhou
PEOPLE'S REPUBLIC OF CHINA

Name of Object: Laboratory Equipment Thermal Cycler

Model(s): TC-XP-A, TC-XP-B, TC-XP-D, TC-XP-F, TC-XP-G

Description of Object:

Rated voltage: 220-240V AC
Rated frequency: 50Hz
Rated input power: 800W
Protection class: I

Tested according to:

EN 61326-1:2006

This EC-Attestation of Compliance is issued according to the Directive 2004/108/EC relating to electromagnetic compatibility on a voluntary basis. It confirms that the listed apparatus complies with all essential requirements of the EMC directive and applies only to the sample and its technical documentation submitted to TÜV SÜD Product Service GmbH for testing and certification. See also notes overleaf.

Test report no.: 70888404501-01

Date, 2009-05-27

Linda Brin



After preparation of the necessary technical documentation as well as the conformity declaration the required CE marking can be affixed on the product. Other relevant directives have to be observed.

Page 1 of 1



Product Service

Attestation of Conformity

No. N8 13 05 48128 046

Holder of Certificate: Hangzhou Bioer Technology Co., Ltd.

1192BinAn Rd, Binjiang District
310053 Hangzhou
PEOPLE'S REPUBLIC OF CHINA

Product: Laboratory Equipment
Thermal cycler

Model(s): TC-XP-A, TC-XP-B, TC-XP-D, TC-XP-F, TC-XP-G
For model differences refer to Page 2

Parameters:

Rated voltage:	220-240VAC
Rated frequency:	50-60Hz
Rated input power:	800W MAX
Protection class:	Class I
Degree of protection:	IP20

Tested according to: EN 61010-1:2010

This Attestation of Conformity is issued on a voluntary basis according to the Low Voltage Directive 2006/95/EC relating to electrical equipment designed for use within certain voltage limits. It confirms that the listed equipment complies with the principal protection requirements of the directive. It refers only to the particular sample submitted for testing and certification. See also notes overleaf.

Test report no.: 70316404501-01

Date, 2013-05-29


(Junjie Sun)



CE After preparation of the necessary technical documentation as well as the EC conformity declaration the required CE marking can be affixed on the product. Other relevant directives have to be observed.

Page 1 of 2

Attestation of Conformity
No. N8 13 05 48128 046



Product Service

The following models will be covered by above certificate:
 TC-XP-A, TC-XP-B, TC-XP-D, TC-XP-F, TC-XP-G

The electrical construction and mechanical construction of above are similar.

The blocks of equipment are different:

Model	Sample Number	Test-tube Volume	Maximum Gradient(°C)	Minimum Gradient(°C)	Block
TC-XP-A	96	0.2ml	--	--	Single
TC-XP-B	60	0.5ml	--	--	Single
TC-XP-D	48x2	0.2mlx2	--	--	Dual
TC-XP-F	384wells block		--	--	Single
TC-XP-G	96	0.2ml	30	1	Gradient