



Product Service

EC-Attestation of Conformity

No. N8 05 04 48128 012

Holder of Certificate: Hangzhou Bioer Technology Co., Ltd.

1192BinAn Rd,Binjiang District
310053 Hangzhou
PEOPLE'S REPUBLIC OF CHINA

Product: Laboratory equipment
Cooling & Heating Block

This EC-Attestation of Conformity is issued on a voluntary basis according to the Low Voltage Directive 73/23/EEC relating to electrical equipment designed for use within certain voltage limits. It confirms that the listed equipment complies with the principal protection requirements of the directive. It refers only to the particular sample submitted for testing and certification. See also notes overleaf.

Test report no.: 70316404503



Date, 2005-05-09

CE After preparation of the necessary technical documentation as well as the conformity declaration the required CE marking can be affixed on the product. Other relevant directives have to be observed.

Page 1 of 2



Product Service

EC-Attestation of Compliance

No. E8N 09 01 48128 024

Holder of Certificate: Hangzhou Bioer Technology Co., Ltd.

1192BinAn Rd,Binjiang District
310053 Hangzhou
PEOPLE'S REPUBLIC OF CHINA

Name of Object: Laboratory Equipment Cooling & Heating Block

Model(s): CHB-202, HB-202

Description of Object:	Rated voltage:	AC 220-240V
	Rated frequency:	50Hz
	Rated power:	120W
	Protection class:	I

Tested according to: EN 61326-1:2006

This EC-Attestation of Compliance is issued according to the Directive 2004/108/EC relating to electromagnetic compatibility on a voluntary basis. It confirms that the listed apparatus complies with all essential requirements of the EMC directive and applies only to the sample and its technical documentation submitted to TÜV SÜD Product Service GmbH for testing and certification. See also notes overleaf.

Test report no.: 70888404503-01

Date, 2009-01-20



CE After preparation of the necessary technical documentation as well as the conformity declaration the required CE marking can be affixed on the product. Other relevant directives have to be observed.



Product Service

EC-Attestation of Conformity
No. N8 05 04 48128 012

Model(s): CHB-202, HB-202

Parameters:

Rated voltage: 220-240 VAC
Rated frequency: 50/60Hz
Rated Input Power: 120W
Protection class: I
Degree of protection: IP20

The electrical construction and mechanical construction of these two models are similar except the function of equipment.

MODEL	FUNCTION
CHB-202	cooling and heating
HB-202	only heating

Tested according to:

EN 61010-2-010:2003