

# Verification of Conformity



Certificates No: **SCT20110625Rocker**

Report Reference No: **08.09.20.0625.01**

Certificate's Holder: **Rocker Scientific Co.,Ltd**

Address: **19F-6, No.206, Guanghua 1st Rd., Lingya Dist.,Kaohsiung City 802, Taiwan**

Product: **Electric Sterilizer**

Model(s): **Dragon320, Dragon300**

Rating(s): **180-240V~, 50Hz, Class I, 220W**

Directives: **Low Voltage Directive 2014/35/EU**

Standards: **EN 61010-1:2010+A1:2019**

**Remark:** This Verification of Compliance has been issued on a voluntary basis. SCT confirms that a Technical Construction File (TCF) is existent for the above listed product(s). The TCF satisfactorily covers the essential requirements of the above listed Directive(s).

Other relevant Directives have to be observed in case they are applicable.

This Document is only valid for the equipment and configuration described and in conjunction with the TCF detailed above. Whereas the Manufacturer is responsible of the certification of the product(s) and not exempted to perform all the necessary activities before placing the product(s) on the market.

The Manufacturer is also responsible of the internal production control to ensure the product(s) are in compliance with the essential requirements of the above mentioned Directive(s).

2020-11-21

Issued Date

