**DATASHEET** | **PART NUMBER** 923050

# Wide open end serological pipette 2ml - CLEARLine® sterile, free from DNase, RNase, and Endotoxins




- Volume : 2 ml
- Div 1/50
- Negative graduation : -2 ml
- Color code : green
- Outer Diameter : 6.1±0.2 mm
- Total Length : 230±1 mm

- Open End serological pipettes in transparent polystyrene
- Specially designed pipettes with an open-end tip, suitable for use with high-viscosity liquids. They ensure high-speed filling and dispensing, and are recommended for applications involving oils, cosmetics, paints, sludge, food products, and cell cultures.
- Individually packaged in paper and plastic film.
- Sterility - The product has been sterilized in accordance with the requirements of ISO 11137, The products meet a sterility assurance level (SAL) of 10-6.
- Free of Bacterial Endotoxins - Tested and met the criteria set forth in the current version of ANSI/AAMI ST 72, "Bacterial Endotoxin Test Methodologies, Routine Monitoring, and Alternatives to Batch Testing."
- BSE/TSE - This product is considered Animal Free as it does not contain any animal-derived materials.
- DNase/RNase Free : Tested by nuclease analysis and found free of detectable DNase/RNase contamination. .
- Volumetric Accuracy - Serological pipettes have an accuracy of ≤ 2% at full volume in accordance with ASTM E934, and ISO 12771
- Food Contact - Raw material complies with EC Regulation No. 1935/2004
- Filter - All pipettes are assembled with a high-quality filter.
- Graduation Ink - Resistant to contact with strong acids, strong bases and strong oxidants.

**Units : 800**

Dominique DUTSCHER SAS - 2C, rue de Bruxelles - 67170 BERNOLSHEIM
Tél : +33 (0)3 88 59 33 90 Fax : +33 (0)3 88 59 33 99 Mail : info@dutscher.com Web : www.dutscher.com