LADR GmbH MVZ Bremen
Bereich Lebensmittel- und Umweltanalytik
Friedrich-Karl-Str. 22, D-28205 Bremen

ISO 17025 accredited by

DAkkS
Deutsche
Akkreditierungsstelle
D-PL-13183-01-00

Test Report
– Eppendorf Consumables –

1 Set Conical Tubes 15 mL

BioBased, screw cap, sterile, pyrogen-, DNase-, RNase-, and DNA-free, 400 Tubes, (8 packs x 50 Tubes)

REF 0030 122 631
GTIN 04043758045653
LOT P225917H

2031-02-10

Read the instructions for use on www.eppendorf.com/manuals
U.S. Pat. www.eppendorf.com/ip

STERILE R

(01) 04043758045653
(10) P225917H
(11) 260210
(17) 310210
(240) 0030122631
(30) 1

Eppendorf SE
Barkhausenweg 1
22339 Hamburg · Germany
eppendorf@eppendorf.com
www.eppendorf.com

Distributed in North America by
Eppendorf North America, Inc. · USA
info@eppendorf.com
www.eppendorfna.com

Made in Germany

eppendorf

- **Human DNA**
  Method: *real-time* PCR acc. to SOP 99880/V8
  Limit: < 2 pg; less than one human cell
- **Bacterial DNA**
  Method: *real-time* PCR acc. to SOP 100567/V5
  Limit: < 50 fg; less than 10 *E. coli* cells
- **DNase**
  Method: DNA Digestion acc. to SOP 100570/V4
  Limit: Not detectable
  (LOD: 1.0 x 10^{-7} Kunitz units)
- **RNase**
  Method: RNA Digestion acc. to SOP 100571/V4
  Limit: Not detectable
  (LOD: 1.0 x 10^{-9} Kunitz units)
- **Sterility**
  Method: Sterilization with SAL 10^{-6} by
  R: Irradiation according DIN EN ISO 11137
  Test method: USP, Ph. Eur. 10.0 / 2.6.12
- **Endotoxin**
  Method: Kinetic-chromogenic LAL test (Ph. Eur. 10.0 / 2.6.14)
  Limit: < 0.001 EU/mL

Customer: Eppendorf SE, Barkhausenweg 1, D-22339 Hamburg

Date of sample receipt / examination: 17.03.2026

Testing of the rinse solution of the above described lot showed conformity within the limits of detection.

24.03.2026
Date of release

Epp 2603-7406
Registration number

i.V. P. Kuhnert
Dr. rer. nat. Sybille Kuhnert (Head)

The test results refer exclusively to the items tested as arrived. The test report must not be copied partially without the approval of LADR GmbH.

LADR GmbH MVZ Bremen, Bereich Lebensmittel- & Umweltanalytik | Friedrich-Karl-Str. 22 | D-28205 Bremen | Germany
Phone: +49 421 4307-500 | Fax: +49 421 4307-199 | **molekularbiologie-bremen@ladr.de** | www.ladr.de

0012 585.730-08