Thermo Fisher Scientific
Water Analysis Instruments
275 Aiken Road, Asheville, NC 28804 USA
Phone: 1-866-984-3766
Email: WLP.Techsupport@thermofisher.com
ThermoFisher.com

Certificate of Analysis

Product Information

Part Number:	951213
Description:	ORION AMMONIA ISE STORAGE SOLUTION 475ML
Lot Number:	352379
Expiration Date:	2032-06-30
Method of Analysis:	Potentiometric

Actual Analysis	Acceptance Range	Results
Ammonium Chloride Concentration	0.078 - 0.088 M	0.082 M

Traceability

NIST Chemical	
NIST Standard: Name	Potassium Chloride
NIST Standard: SRM#	999c

Certified By:



Thermo Fisher Scientific
Water Analysis Instruments (formerly Water and Lab Products)
Quality Assurance
Certified Date: 2026-06-23

DOMINIQUE DUTSCHER SAS

At Thermo Fisher Scientific, our mission is to enable our customers to make the world healthier, cleaner, and safer. To support this mission, where applicable as listed on this document, our manufacturing site utilizes NIST (National Institute of Standards and Technology) certified reference standards during the testing. Additionally, our testing equipment is calibrated and traceable to NIST standards, ensuring precise and consistent results.

For Laboratory Use. It is the customer's responsibility to ensure that the performance of the product is suitable for customer specific use or application. ©2025 Thermo Fisher Scientific Inc. All rights reserved. All trademarks are the property of Thermo Fisher Scientific and its subsidiaries unless otherwise specified.