



DECLARATION / DECLARATION

Fabricant / Manufacturer :

STERLAB

2720 chemin Saint-Bernard

06224 VALLAURIS CEDEX - FRANCE

Tel : 04.97.24.58.58 – Fax : 04.97.24.58.59 – e-mail : contact@sterlab.com

Nous garantissons et déclarons que les indicateurs chimiques STERIPOINT® :

- sont conformes à la norme **ISO 11140 (2009)** (anciennement série EN 867) relative aux indicateurs de stérilisation vapeur et à la norme **EN 285 (2006) + A2 (2009)** relative aux grands stérilisateur vapeur
- sont des **dispositifs médicaux de classe I** au sens de la directive 93/42/CEE (voir déclaration CE de conformité)
- sont des **indicateurs chimiques de classe 6** selon la norme EN ISO 11140-1 (anciennement Classe D selon la norme EN 867-1).

We assure and certify that the chemical indicators **STERIPOINT® :**

- comply with ISO 11140 standard (2009) (previous EN 867) relative to steam sterilisation indicators and with **EN 285 (2006) + A2 (2009)** standard relative to large steam sterilizers
- are **medical devices class I** according to Directive 93/42/EEC (see EC Conformity Declaration)
- are **chemical steam indicators class 6** according ISO 11140-1 standard (previous class D according to EN 867-1).

Etabli le : 12/01/2015

Issued on : 12/01/2015

Nom / Name : Paul MARTEL

Fonction / Function : Président

Signature :

STERLAB
2720 Chemin St-Bernard
06224 VALLAURIS Cedex - FRANCE
Tél. : (33) 04 97 24 58 58
Fax : (33) 04 97 24 58 59