

# EC- DECLARATION OF CONFORMITY

RE: Medline Scientific Heating and Stirring Apparatus

Models Covered:

- 'MS' / Hotplates & Stirrers
- 'E' / Controlled Heating Mantles
- 'EAM' / Multi-position Heating Mantles
- 'ES' / Combined Stirrer / Heating Mantles

Dear Sir / Madam,

We herewith certify our appliances to be built and tested according to approved test procedures and comply fully with the respective standards.

## **1. Electrical safety and CE mark**

Complies with the essential requirements of the following directives EMC Directive 2004/108/EC Low Voltage Equipment Directive 2006/95/EC

Standards and technical specifications referred to:

BS EN 61326-1: 2013  
BS EN 61000-3-2: 2014  
BS EN 61000-3-3: 2013  
BS EN 55011/A1: 2010 (Class A)  
BS EN 61010-1: 2010

## **2. Test Procedures**

Quality control according to the technical handbook on all

- Optical parts
- Mechanical parts
- Electrical parts

All above goods are classified to be of UK preferential origin.

Medline Scientific Ltd

Chalgrove, UK

August 2020