



**Bellco Glass, Inc.**  
340 Edrudo Road  
Vineland, NJ 08360



Contact: (856) 691-1073 (800) 257-7043 Fax: (856) 691-3247  
E-Mail: [cservice@bellcoglass.com](mailto:cservice@bellcoglass.com) Internet: <http://www.bellcoglass.com>

## CERTIFICATE OF COMPLIANCE

Date: 2/9/24

This is to certify that the items referenced below have been manufactured and inspected in compliance with all applicable Bellco Glass, Inc. material and process specifications and are constructed of these materials. Bellco Glass, Inc. declares that our products are Dioxin and melamine free. All products produced by Bellco Glass, Inc. are free from Transmissible Spongiform Encephalopathy (TSE) and Bovine Spongiform Encephalopathy (BSE).

<b>Catalog#:</b>	<b>Description</b>	<b>Material</b>
A523-600	White silicone gasket	Silicone

1. Type I, **Class A** Borosilicate clear glass conforming to federal specification DD-G-54(1)(b) and ASTM E-438 and meets the USP specifications for Type I Class A Borosilicate Clear glass.
2. Type 1, **Class B** Borosilicate clear glass (N-51A) conforming to ASTM E-438 and meets all USP specifications for Type 1 Class B Borosilicate clear glass.
3. The Polypropylene is constructed of materials which conform to FDA specification 21CFR 177.1520 (Olefin Polymers). Colorant components that are FDA-regulated comply with 21CFR 177.3297 (Colorants for Polymers).
4. The ingredients used to construct the silicone rubber comply with FDA 21CFR 177.2600 and meets the requirements for USP Class VI.
5. The stainless-steel materials used comply with AISI 316/316L Stainless Steel.
6. The PTFE is constructed of materials which conform to FDA 21CFR 177.1550, USP Class V1 material.
7. Delrin (Acetal resin) is constructed of materials which conform to FDA specification 21CFR177.2480
8. Lexan Polycarbonate Resin is constructed of materials which conform to FDA specification 21CFR 177.1640 and meets the requirements for USP Class VI.
9. Polystyrene is constructed of materials which conform to FDA 21CFR 177.1640.
10. The PBT is constructed of materials which conform to FDA 21CFR 177.1660.
11. The aluminum material complies with ASTM B209 and ASMESB209.
12. The EPDM Rubber conforms to FDA 21CFR 177.2600 Rubber Articles intended for repeated use.
13. The neoprene Rubber conforms to FDA 21CFR 177.2600 Rubber Articles intended for repeated use.
14. The Cotton Roving conforms to FDA 29CFR 1910.262
15. The Polyurethane complies with FDA 21CFR. 177.1680

*Stephen Agliano*  
**Quality Assurance**