

## Certificate of Analysis

### CERTIFIED REFERENCE MATERIAL

Solution of Yttrium(Y) concentration 10000 mg/l Matrix: 5% HNO<sub>3</sub>

Lot N: CXXXXXX  
Barcode: XXXXXXXX

Ref N: C167.5NP.L1

Certification Date:XXXXXX

Component	Certified Value and uncertainty [mg/l]	Metrological traceability
Y	10032 ± 30 <sup>(y)</sup>	NIST SRM No 3167a Lot 120314; NIST SRM No 3168a Lot 120629

**Notes:**

(y) WQP 5.15.1.24 The certified value was obtained by a weighted mean of the results of two independent calibration methods among: Classical Volumetric, Primary Gravimetric, Instrumental (ICP/OES, ICP/MS or IC)

Density\* 1.055 g/cm<sup>3</sup> at 20°C

Starting Material, Purity*	Batch
Y <sub>2</sub> O <sub>3</sub> 99.95%	82110598

\* These values are not certified

Storage Conditions: Store under normal laboratory conditions, at temperatures between 15° to 25°C

Shelf-life: XXXXXXXXXX

Date of opening: .....

(Recommended period of use should not exceed 12 months from date of opening)

**Concept of Certification and traceability statement:**

This certified reference material is produced using a high purity starting material, acid from sub-boiling and 18 MOhm deionized water. The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k = 2$ , which for a normal distribution corresponds to a coverage probability of approximately 95%. The standard uncertainty of measurement has been determined in accordance with EA 4/02

Property of the result of a measurement whereby it can be related to stated references, usually national or international standards, through an unbroken chain of comparisons all having stated uncertainties (ISO VIM)

The metrological traceability is assured through calibration on ICP/OES. The calibration curve is drawn using a series of standard solutions prepared from a certified reference material traceable to SI of NIST (SRM) or BAM (CRM). All contributions in relation to the certification of standard solutions are considered when evaluating the uncertainty.

The measurement results are traceable to SI. All analytical balances used for the preparation of the solution are calibrated yearly under an in-house procedure with analytical weights, traceable to DKD, and are checked daily. Class A laboratory glassware is used.

The results from temperature measurement are traceable to SI. The thermometers used for solution's calibration are calibrated from an ISO 17025 accredited laboratory. The ambient conditions are controlled with a hygrometer calibrated from an ISO 17025 accredited laboratory.

**Intended use: For Laboratory Use Only**

Calibration of ICP/OES, AAS  
Preparation of "working reference samples"  
This statement is not intended to restrict the use for other purposes.

Validation of analytical methods  
Detection limit and linearity studies

**Instructions for the correct use of this reference material:**

This certified reference material can be used directly or can be diluted in an appropriate high purity matrix. Only a clean class A glassware should be used. Do not pipet from container. Obtained concentration (in mg/l) after dilution is a result from the multiplication of certified value of CRM concentration and the CRM's volume used for dilution and divided into the flask's volume used for dilution.

**Stability and storage:**



This CRM is with a guaranteed stability until  $\pm 0.5\%$  of the certified concentration within its shelf life. Stability is guaranteed, provided that the solution is kept in its original packaging, tightly closed stored, as written in the section: Storage Conditions..

**Hazardous situation:**

The normal laboratory safety precautions should be observed when working with this CRM. Further details for the handling of this CRM are available as safety data sheet.

**Level of homogeneity:**

The material was tested for homogeneity by analyzing randomly selected samples according to an in-house procedure. The level of homogeneity proved satisfactory for a sample volume of 20 ml. The uncertainty incorporates the sample standard deviation combined with the uncertainty calculated from homogeneity and stability studies.

To ensure sufficient homogeneity of the sample prior to use thoroughly mix by inversion.

*Responsible for quality control: B. COULANGE\**

*\*This certificate has been computer generated and does not signed*

*This document QF 5.17.1/1 version 1 is designed and the certified value(s) and uncertainty(ies) are determined in accordance with ISO Guide 31, ISO Guide 35, and Eurachem / CITAC Guides*

*This certificate relates solely to the lot number given above.*

*All processes (including generating of this certificate) are completely controlled by the specialized Computer-Aided-Manufacturing (CAM) software.*

*This Certified Reference Material was produced under a quality management system that is:*

*- Registered to ISO 9001 Quality Management System (Lloyd's Register Quality Assurance Ltd Cert No 0039638)*

*- Accredited according to ISO/IEC 17025 – Testing (ANAB Cert No AT-1836)*

*- Accredited according to ISO 17034 - Reference Material Producer (ANAB Cert No AR-1835)*

**Trace impurities in the actual solution reported in ppm:**

(all values below are nominal and not certified)

Ag	<0.0038
Al	<0.0018
As	<0.016
Au	<0.016
B	0.95
Ba	<0.0005
Be	<0.0001
Bi	<0.016
Ca	<0.004
Cd	<0.0012
Ce	<0.0085
Co	<0.0028
Cr	<0.0014
Cs	<0.05

Cu	<0.0009
Dy	<0.0054
Er	<0.0035
Eu	<0.0039
Fe	<0.0025
Ga	<0.020
Gd	<0.0028
Ge	<0.020
Hf	<0.0032
Hg	<0.024
Ho	<0.0053
In	<0.098
Ir	<0.0061
K	<0.0095

La	<0.0024
Li	<0.0001
Lu	<0.0062
Mg	0.3
Mn	<0.001
Mo	<0.0024
Na	0.2
Nb	<0.0066
Nd	<0.0058
Ni	<0.0061
P	<0.048
Pb	<0.021
Pd	<0.033
Pr	<0.0046

Pt	<0.0097
Rb	<0.063
Re	<0.0081
Rh	<0.0038
Ru	<0.0089
S	<0.071
Sb	<0.020
Sc	<0.0016
Se	<0.023
Si	2
Sm	<0.0058
Sn	<0.050
Sr	<0.00006
Ta	<0.004

Tb	<0.022
Te	<0.031
Th	<0.014
Ti	<0.0012
Tl	<0.028
Tm	<0.0023
U	<0.45
V	<0.0018
W	<0.017
Y	*
Yb	<0.0003
Zn	<0.0032
Zr	<0.0007