

# Standard Analysis Certificate

**PRODUCT** : BUFFER SOLUTION pH 9 +/-0.02 BLUE  
**CODE** : PS0427.

TEST	U.M.	SPECIFICATION
pH	unité pH	8.98 - 9.02
Temperature of measurement	°C	19 - 21

Accuracy : +/- 0.02 pH units - Standardized against NIST reference materials.

Date

:29/02/2020

**QUALITY CONTROL RESPONSIBLE**  
**P. GUERARD (PEYPIN)**