CPA chem
The experts in custom-made Certified Reference Materials

# Certificate of Analysis

## CERTIFIED REFERENCE MATERIAL

**TIC Standard Solution 1000mg/l as sodium carbonate (standard solution acc. to EN 1484)**

Lot N: XXXXXX Ref N: TIC1000.L5 Certification Date:XXXXXX
Barcode: XXXXXXXX

| Component | Certified Value and uncertainty [mg/l] | Metrological traceability |
|---|---|---|
| C | 1000.1 ± 3.0 (y) | BAM CRM No 93440 Lot BCBM3204V |

Notes:
(y) WQP 5.15.1.24 *The certified value was obtained by a weighted mean of the results of two independent calibration methods among: Classical Volumetric,Primary Gravimetric, Instrumental (ICP/OES, ICP/MS or IC)*

Density* 1.005 g/cm$^3$ at 20°C

| Starting Material, Purity* | Batch |
|---|---|
| $Na_2CO_3$ 99.995% | 82105914 |

* *These values are not certified*

Storage Conditions: Store under normal laboratory conditions, at temperatures between 15° to 25°C

Shelf-life: XXXXXXXXXX Date of opening: ……..

*(Recommended period of use should not exceed 1 months from date of opening)*

**Concept of Certification and traceability statement:**
*This certified reference material is produced using a high purity starting material, acid from sub-boiling and 18 MOhm deionized water.*
*The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k = 2, which for a normal distribution corresponds to a coverage probability of approximately 95%. The standard uncertainty of measurement has been determined in accordance with EA 4/02*
*Property of the result of a measurement whereby it can be related to stated references, usually national or international standards, through an unbroken chain of comparisons all having stated uncertainties (ISO VIM)*
*The metrological traceability is assured through calibration on . The calibration curve is drawn using a series of standard solutions prepared from a certified reference material traceable to SI of NIST (SRM) or BAM (CRM). All contributions in relation to the certification of standard solutions are considered when evaluating the uncertainty.*
*The measurement results are traceable to SI. All analytical balances used for the preparation of the solution are calibrated yearly under an in-house procedure with analytical weights, traceable to DKD, and are checked daily.*
*Class A laboratory glassware is used.*
*The results from temperature measurement are traceable to SI. The thermometers used for solution's calibration are calibrated from an ISO 17025 accredited laboratory. The ambient conditions are controlled with a hygrometer calibrated from an ISO 17025 accredited laboratory.*

**Intended use: *For Laboratory Use Only***
Calibration of
Preparation of "working reference samples"
Validation of analytical methods
Detection limit and linearity studies
This statement is not intended to restrict the use for other purposes.

**Instructions for the correct use of this reference material:**
This certified reference material can be used directly or can be diluted in an appropriate high purity matrix. Only a clean class A glassware should be used. Do not pipet from container. Obtained concentration (in mg/l) after dilution is a result from the multiplication of certified value of CRM concentration and the CRM's volume used for dilution and divided into the flask's volume used for dilution.

**Stability and storage:**
This CRM is with a guaranteed stability until ±0.5% of the certified concentration within its shelf life. Stability is guaranteed, provided that the solution is kept in its original packaging, tightly closed stored, as written in the section: Storage Conditions.The laboratory performs stability tests according to MQP 5.14.1

PROFORMA DUTSCHER SAS



CPAchem Ltd
www.cpachem.com
e-mail: info@cpachem.com; tel.:+359 42 60 77 16
for France, Belgium and Switzerland:
e-mail: acsd2@wanadoo.fr; tel.:01 30 57 57 32 / fax: 01 30 57 57 33

therefore solutions with one and the same bar-code number might have different expiration dates.

**Hazardous situation:**
The normal laboratory safety precautions should be observed when working with this CRM. Further details for the handling of this CRM are available as safety data sheet.

**Level of homogeneity:**
This solution was mixed according to an in-house procedure and is guaranteed to be homogeneous.
To ensure sufficient homogeneity of the sample prior to use thoroughly mix by inversion.

**Names of certifying officers:**

Laboratory:  Tihomir Stoyanov

Manager:  Krassimira Taralova

*This document QF 5.17.1/1 version 1 is designed and the certified value(s) and uncertainty(ies) are determined in accordance with ISO Guide 31, ISO Guide 35,and Eurachem / CITAC Guides*

*This certificate relates solely to the lot number given above.*
*All processes (including generating of this certificate) are completely controlled by the specialized Computer-Aided-Manufacturing (CAM) software.*

*This Certified Reference Material was produced under a quality management system that is:*
*- Registered to ISO 9001 Quality Management System (Lloyd's Register Quality Assurance Ltd Cert No 0039638)*
*- Accredited according to ISO/IEC 17025 – Testing (ANAB Cert No AT-1836)*
*- Accredited according to ISO 17034 - Reference Material Producer (ANAB Cert No AR-1835)*

DOMINIQUE DUTSCHER SAS